

# CCQM-K84 Final Report

## International comparison CCQM-K84

### Carbon monoxide in Synthetic air at ambient level

Jeongsoon Lee<sup>1</sup>, Dongmin Moon<sup>1</sup>, Jinbok Lee<sup>1</sup>, Jeongsik Lim<sup>1</sup>, Brad Hall<sup>2</sup>, Paul Novelli<sup>2</sup>, Paul J. Brewer<sup>3</sup>, Michael Miller<sup>3</sup>, Arul Murugun<sup>3</sup>, Marta Doval Minarro<sup>3</sup>, Han Qiao<sup>4</sup>, Hu Shuguo<sup>4</sup>, L. A. Konopelko<sup>5</sup>, Y.A. Kustikov<sup>5</sup>, A.V. Kolobova<sup>5</sup>, V.V. Pankratov<sup>5</sup>, I.I. Wasserman<sup>5</sup>, S.V. Zav'yalov<sup>5</sup>, O.V. Efremova<sup>5</sup>, M.V. Pavlov<sup>5</sup>, Gerald Mitchell<sup>6</sup>, Frank Guenther<sup>6</sup>, Jari Walden<sup>7</sup>; Nobuyuki Aoki<sup>8</sup>, Takuya Shimosaka<sup>8</sup>, Mace Tatiana<sup>9</sup>, F. Lagler<sup>10</sup>, Annette Borowiak<sup>10</sup>

<sup>1</sup>Korea Research Institute of Standards and Science (KRISS), Division of Metrology for Quality Life, P.O.Box 102, Yusong, Daejeon, Republic of Korea

<sup>2</sup> National Oceanographic and Atmospheric Administration, Boulder, Colorado, United States of America, 325 Broadway Boulder, Colorado 80305, USA

<sup>3</sup> National Physical Laboratory (NPL), Teddington, Middlesex, TW11 0LW, UK

<sup>4</sup> National Institute of Metrology(NIM), 18 Beisanhuan Donglu, Beijing 100013, China

<sup>5</sup> D.I. Mendeleev Institute for Metrology, Rostekhnregulirovaniye of Russia (VNIIM), Department of State Standards in the field of Physical Chemical Measurements, 19, Moskovsky Prospekt, 198005 St-Petersburg, Russia

<sup>6</sup> National Institute of Standards and Technology (NIST), Chemical Science and Technology Laboratory, 100 Bureau Drive, Gaithersburg MD, USA

<sup>7</sup> Finnish Meteorological Institute (FMI), Air Quality Research P.O. Box 503, FI-00101 HELSINKI, Finland

<sup>8</sup> National Metrology Institute of Japan (NMIJ), Tsukuba Central 3, Tsukuba 305-8563, Japan

<sup>9</sup> Laboratoire national de metrologie et d'essais (LNE), rue Gaston Boissier - 75724 Paris Cedex 15, France

<sup>10</sup> Joint Research Center, Institute for Environment and Sustainability Via E. Fermi 2749, 21027 Ispra, Italy

Errata: With Table 10 included

---

1 **Field**

2 Amount of substance

3 **Subject**

4 Carbon monoxide in Synthetic air at ambient level

5

6 **Participants**

7 KRISS, NIST, LNE, NIM, NMIJ, VNIIM, FMI, JRC, NOAA, NPL

8

9 **Organizing body**

10 CCQM GAWG

11

12

1	
2	Table of Contents
3	
4	Introduction
5	Supported claims
6	Participants
7	Schedule
8	Comparison protocol
9	Gravimetric preparation
10	CO Stability in air
11	Measurement results
12	DoE
13	Conclusions
14	References
15	Appendix: Measurement Reports
16	FMI
17	JRC
18	KRISS
19	LNE
20	NIM
21	NIST
22	NMIJ
23	NOAA
24	NPL
25	VNIIM
26	

## 1 **Introduction**

2 Carbon monoxide (CO) is reported to be mainly emitted from industries, transportation, and  
3 burnings for various usages. Its atmospheric lifetime varies from weeks to months, depending  
4 on the mixing ratio of the highly reactive hydroxyl radical. Even though the ambient level of  
5 CO varies as a function of regional sources, the mixing ratio of CO ranges from 30 nmol/mol  
6 to 300 nmol/mol at the marine boundary layers and from 100 nmol/mol to more than 500  
7 nmol/mol in urban areas<sup>(1)</sup>. In order to study temporal trends and regional variation of the  
8 level of CO, the National Oceanic & Atmospheric Administration/Earth System Research  
9 Laboratory-Global Monitoring Division (NOAA/ESRL-GMD<sup>(2)</sup>) has played a key role as the  
10 designated Central Calibration Laboratory (CCL) within the frame of the World  
11 Meteorological Organization (WMO) Global Atmosphere Watch (GAW) program.  
12 NOAA/ESRL-GMD provides natural air standards, analyzed for CO, to WMO GAW  
13 participants. Since the structure of WMO traceability chain appears hierarchical and explicit  
14 all over the world, WMO intends to improve the CO measurement compatibility to up to 2  
15 ppb (in case of extensive compatibility goal: 5 ppb, GAW report No. 213<sup>(3)</sup>) in order to ensure  
16 compatibility through the GAW network. Nevertheless, accurate measurement of CO at  
17 ambient level has been proven difficult due to the lack of stability in cylinders. For these  
18 reasons, it is necessary that measured results are compared among the values assigned by  
19 various NMIs.

20 This key comparison was initially proposed to aim at a CO/N<sub>2</sub> standard in the 2010 CCQM  
21 meeting by KRISS. With participation of FMI, NOAA, and Empa, a modified scheme of  
22 CO/air standards was developed for the purpose of atmospheric observations and co-  
23 operative support to WMO/GAW activities. Therefore, the purpose of the comparison is to  
24 support the measurement capability of CO at ambient level of 350 nmol/mol. Further, this  
25 key comparison is expected to contribute to the establishment of traceability to a single scale  
26 of CO between NMIs by means of harmonizing the results from different national standards.  
27 The Empa result lies in a different report.

## 1 **Supported claims**

2 This key comparison support the measurement capability, which can be used to support CMC  
3 claims, for carbon monoxide in air, synthetic air, and nitrogen from 50 nmol/mol to 1000  
4 nmol/mol.

## 6 **Participants**

7 A total of ten CCQM members took part in this key. Among them, NOAA, as a WMO  
8 designated lab for CO, participated under MRA. The participants are listed in Table 1.

9  
10 Table 1: List of participants

Acronym	Country	Institute
FMI*	FI	Finnish Meteorological Institute
JRC	IT	Joint Research Center
KRISS	KR	Korea Research Institute of Standards and Science, Daejeon, Republic of Korea
LNE	FR	Laboratoire National d'Essais, Paris, France
NIM	CN	National Institute of Metrology, Beijing, P.R.China
NIST	US	National Institute of Standards and Technology, Gaithersburg, United States of America
NMIJ	JP	National Metrology Institute of Japan, Tsukuba, Japan
NOAA	US, (CCL/WMO)**	National Oceanographic and Atmospheric Administration, Boulder, Colorado, United States of America
NPL	UK	National Physical Laboratory, Teddington, Middlesex, United Kingdom
VNIIM	RU	D.I. Mendeleev Institute for Metrology, St. Petersburg, Russia

1 \* Standard from the participating laboratory; these laboratories do not make use of their own standards.

2 \*\*CCL/WMO: Central Calibration Laboratory/ World Meteorological Organization

3

## 4 **Schedule**

5

6 The Schedule for this key comparison was as follows:

7

Date	
Mar. , 2012	Preparation/verification of mixtures by KRISS
Apr. , 2012	Registration and protocol circulation
Until July , 2012	Shipment of cylinders from KRISS to participants
Until Aug. , 2012	Measurement by participants and sending report to KRISS
Until Mar. , 2013	Return of cylinder to KRISS
Until May, 2013	Second verification for returned cylinders
Until Nov. , 2013	3 <sup>rd</sup> verification
Until Mar. , 2014	4 <sup>th</sup> verification and Draft A report
Until Nov. , 2014	Draft B report

## 8 **Comparison protocol**

9 A set of mixtures of carbon monoxide in synthetic air of the nominal mole fraction of  
10 approximately 350 nmol/mol, were gravimetrically<sup>(4)</sup> prepared by the coordinating laboratory  
11 of KRISS. Each mixture was then verified by means of a GC/FID/Methanator system, against  
12 very fresh primary standard gas mixture (PSM) with amount-of-substance fractions of approx.  
13 350 nmol/mol. The amount-of-substance fractions determined by KRISS were adopted as key  
14 comparison reference values (KCRV).

15

16 Table 2. Nominal composition of mixtures

Component	X (nominal value)
Carbon monoxide [nmol/mol]	350
Argon [%mol/mol]	0.93

Oxygen [%mol/mol]	21
Nitrogen [%mol/mol]	balance

1

2 The pressure in each cylinder was approximately 100 bar; cylinders of 10 dm<sup>3</sup> (Al. Luxfer,  
3 UK). The amount-of-substance mole fraction obtained from gravimetry, and purity analysis  
4 of parent gases, were used as reference values. Accordingly, each cylinder was assigned its  
5 own reference value. Impurity analysis was performed using the best analytical method.  
6 Participating laboratories were requested to specify in detail which analytical method(s),  
7 which standards were used, and how the evaluation of measurement uncertainty was  
8 performed.

9

10 Each participating laboratory was responsible for the calibration of its own instrument(s) used  
11 for the analysis. Applied calibration method should be well established and reported to  
12 KRISS. This is an absolute necessity for proper evaluation of the data.

13

14 Each laboratory was required to express the uncertainty on all results submitted, as expanded  
15 uncertainty. The evaluation of the measurement uncertainty should be in accordance to the  
16 “Guide to the expression of uncertainty in measurement” (ISO GUM). The participants  
17 should provide a detailed description of the uncertainty budget, including;

18

- 19 - Method of evaluation (type A or type B)
- 20 - Assumed Probability distribution
- 21 - Standard uncertainties and sensitivity coefficients
- 22 - Effective degrees of equivalence (if applicable/used)
- 23 - Statistical reasoning behind the coverage factor

24

25 After the measurements, the participants were instructed to return the cylinders with a  
26 sufficient amount of gas (pressure at least 50 bar) to KRISS for re-analysis.

1  
2 Additional measurement reports and information can be submitted jointly with the report  
3 form to the KRISS, and would be taken into consideration during the evaluation.  
4

## 5 **Gravimetric preparation**


6 A set of carbon monoxide in synthetic air with the nominal mole fraction of approx. 350  
7 nmol/mol, were gravimetrically<sup>(4)</sup> prepared through four step dilutions by the coordinating  
8 laboratory of KRISS. Each mixture was then verified by means of a GC/FID/Methanator  
9 system, against very fresh primary standard gas mixture (PSM) with amount-of-substance  
10 fractions of approx. 350 nmol/mol. After purity analysis of the parent gases and verification  
11 of the prepared mixtures, the amount-of-substance fractions were adopted as key comparison  
12 reference value (KCRV).  
13

14 For purity analysis of O<sub>2</sub>, Ar, and N<sub>2</sub> matrix gases, a GC/FID/Methanator was used and a  
15 detection limit for carbon monoxide was set at 0.7 nmol/mol (3σ). A significant amount-of-  
16 substance fraction of carbon monoxide impurity was measured by 0.7 nmol/mol in nitrogen  
17 and 6.9 nmol/mol in oxygen. Based on this purity results, CO mole fractions are properly  
18 applied to every dilution steps. After shipping, the amount of CO mole fraction changed so  
19 significantly, which must be taken into account for any further analysis. The stability of the  
20 CO in air standards used in this study can be described as follows.  
21

22 A total of 16 cylinders were prepared for the comparison on July 21<sup>st</sup> 2012 and analyzed on  
23 July 24<sup>th</sup> 2012. Because a GC/FID/Methanator shows good linearity over the mole fraction  
24 range tested in this key comparison, a reference cylinder (A) was measured between sample  
25 cylinders as to not only measure sensitivity ( $\text{Sensitivity} = \frac{\text{Response}_{\text{Instrument}}}{\text{Mixing ratio}_{\text{Prepared}}}$ )  
26 ratios of sample versus reference but also correct analyzer drift during the comparison  
27 analysis in a sequence of A-B-A-C-A-..., and so on. Sensitivity ratios between reference and  
28 sample were compared with each other to find that they were within 0.30 % around unity as  
29 shown in Fig. 1. Thus, this 0.30% which satisfies a verification criteria was assigned to the



1 analytical uncertainty ( $k=2$ ) of certified values. In fact, their gravimetric preparation  
 2 uncertainty including purity analysis was 0.22000%,  $k = 2$ . Considering the gravimetric  
 3 preparation and the analytical uncertainty, the total expanded uncertainty of CO cylinder was  
 4 calculated to be 0.37%.



6  
7 Figure 1. Consistency among the prepared cylinders (July 2012)

8  
9 Table 3 lists the set of PSMs including their mole fraction composition. From this set ten  
10 standards were distributed to participants.


11  
12 Table 3. Gravimetric preparation uncertainty of CO gases for a set of cylinders.

Cylinder	CO [nmol/mol]	$U_{\text{prep.}}, k = 2$ [nmol/mol]	Ar [% mol/mol]	O <sub>2</sub> [% mol/mol]	N <sub>2</sub> [% mol/mol]
D015224	356.33	1.33	0.9161	20.72	78.37
D015230	351.52	1.31	0.8986	20.91	78.19
D015283	350.81	1.31	0.9040	21.10	77.99
D015280	350.60	1.30	0.8907	20.96	78.15
D015285	348.86	1.30	0.9035	20.68	78.42
D015215	349.97	1.30	0.9610	20.96	78.08
D015217	350.83	1.31	0.9529	21.71	77.34
D015275	342.44	1.27	0.9362	20.62	78.45
D015286	353.26	1.31	0.9238	20.89	78.19
D015220	352.35	1.31	0.9184	20.67	78.42
D015223	351.35	1.31	0.8983	21.16	77.95

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15

## CO stability in air

In order to verify the mixture's stability, each cylinder was analyzed four times over a 20 month period. Each measurement was carried out against very fresh gravimetric standards aging less than a week, or in case of a reanalysis, a few weeks. First measurement was done, as indicated in the key comparison schedule, before the shipment of the cylinders. Additional three measurements were conducted once the cylinders were returned from the participants. For the second analysis the cylinders were compared to new gravimetric standards produced in March 2013. The second analysis was performed in May 2013. During the second analysis the returned cylinders were compared against the standards prepared in March 2013 (Fig. 2). For the 3th and 4th verifications, the returned cylinders were analyzed in Nov 2013 and Mar 2014 against two newly prepared sets, respectively. The two new sets of standard mixtures had been prepared in Sep. 2013 and Jan 2014, as indicated in figure 2 and table 4.



16  
17 Figure 2. Drift of the cylinders initially prepared for the comparison (from top, filled square: preparation values,  
18 circle, upward triangle and downward triangle: measurement values against three sets of standard mixtures  
19 prepared in the subsequent times, respectively)  
20


21 The second and third analysis showed that CO mole fractions increased by an average of 0.93%  
22 excluding D985730 and D015224 which increased by more than 1.7% for 14 months. Both  
23 Figure 2 and Table 4 show results of consecutive analyses since the preparation of mixtures.  
24 CO appears to have stopped increasing after an eight-month period from the preparation date.  
25 Due to very low amount of substance in D015220, its fourth analysis was skipped. Stability  
26 changes are shown in Figure 3, where the relative differences from the gravimetric values of

1 the cylinders are shown as a function of the analysis time on the x axis. The differences look  
 2 to have been stable (< 0.1 %) since the second verification, excluding one cylinder that  
 3 showed continued drift.

4  
 5 Table 4. Temporal variation of the cylinders for the CCQM-K84

Laboratory	Cylinder	$x_{prep}$	$x_{2nd}$	$x_{3rd}$	$x_{4th}$
		[nmol/mol]	[nmol/mol]	[nmol/mol]	[nmol/mol]
		Jul. 2012	2 <sup>nd</sup> ver. May. 2013	3 <sup>rd</sup> ver. Nov. 2013	4 <sup>th</sup> ver. Mar. 2014
LNE	D015224	356.33	363.65	363.77	366.04
NIST	D015230	351.52	357.62	357.57	357.81
NOAA	D015283	350.81	354.28	354.29	354.19
FMI	D015285	348.86	351.82	352.15	352.29
NPL	D015215	349.97	352.99	353.03	352.72
JRC	D015217	350.83	353.96	354.12	353.96
NMIJ	D015275	342.43	345.48	345.67	345.76
KRISS	D015286	353.26	356.74	356.75	356.80
NIM	D015220	352.35	355.67	355.78	-*
VNIIM	D015223	351.35	354.42	354.70	354.55

6 \*NIM cylinder ran out before 4<sup>th</sup> analysis.  
 7



8  
 9 Figure 3. Mole fraction changes since preparation in July 2012.  
 10

11 In order to validate the long-term stability test described in previous section, a set of CO/N<sub>2</sub>  
 12 cylinders was newly manufactured to be used as a stable reference. It is well known that  
 13 ambient level of CO in N<sub>2</sub> cylinders have been stable over a few years. Our result shows that

CO in N<sub>2</sub> cylinders are stable within 0.02 % even for four years (Table 5).

Table 5. Details on CO stability tests of CO in N<sub>2</sub> cylinders

	$x_{prep}$ [nmol/mol]	Prep. date	Analysis date	Sensitivity*	Matrix
D015249	344.38	2014.1.11	2014.3.08	0.8485	N <sub>2</sub>
D015318	343.35	2014.1.11	2014.3.08	0.8486	N <sub>2</sub>
ME2223	400.18	2010.3.15	2014.3.08	0.8486	N <sub>2</sub>
ME5576	399.75	2010.3.15	2014.3.08	0.8486	N <sub>2</sub>
ME5629	383.25	2010.3.15	2014.3.08	0.8483	N <sub>2</sub>

\* Sensitivity (S) is defined as a ratio of instrumental response (Res) and the gravimetric concentration (X). S=Res/X.

Contrary to Table 5, Table 6 shows that the sensitivities of CO/air cylinders versus that of CO/N<sub>2</sub> vary by 5 %. Since the stability of CO concentrations in N<sub>2</sub> (Table 5) were ensured by the constant sensitivities among 4-year old cylinders (MEXXXX) and newly prepared gravimetric cylinders (DXXXXXX), it is plausible to set new CO/N<sub>2</sub> cylinders as a reference for the validation test (Table 6), of which measurement sequence is R-A-R-B-R-C-R.... In the series measurement ‘R’ denotes a reference measurement and the CO/N<sub>2</sub> cylinder (D015249) was used. Ratios between sensitivities of CO/air and CO/N<sub>2</sub>, namely S(air)/S(N<sub>2</sub>), provides reliable indicator to check the concentration variation within a CO/air cylinder. For instance, the sensitivity ratio value of newly prepared cylinder (D155876/D015249) is lower than that of the older cylinders to exhibit the rate of increase in the CO amounts. As the stability of CO in N<sub>2</sub> cylinders has been proven to be considerably superb (Table 5), a degree of CO-drift among CO/air cylinders can be derived by the comparison of sensitivity ratios of the measurements as follow.

$$RD_{D155876}^{B,C,D,m} (\%) = \left[ \frac{\left| \frac{S(air)}{S(N_2)} \right|_{B,C,D,m}}{\left| \frac{S(air)}{S(N_2)} \right|_{D155876}} - 1 \right] \times 100 \quad (1)$$

where superscript and subscript stands for measurement index. The value of ‘RD’ directly denotes the relative deviation of CO concentration determined against sensitivity ratio of

1 cylinder D155876 to D015249 (CO/N<sub>2</sub>).  
 2 Though the drift rate of CO mole fraction seems to be uneven for every cylinder in Table 6,  
 3 CO mole fraction tends to sit around 1% incensement. The Cylinders of D015224 (LNE) and  
 4 D015230 (NIST) were more enormously varied 1.7 % and 2.7 % than the others. In the long  
 5 run it is the results in table 6 that agree with those in Table 4 and Fig. 3. It can be  
 6 conclusively said that the stability of cylinders for this key comparison was thoroughly  
 7 evaluated and confirmed by the means of CO/N<sub>2</sub> referencing method. In future comparisons,  
 8 the CO stability in air should be monitored for several months prior to distribution.

9

10 Table 6. Stability of CO/air against CO/N<sub>2</sub>

	$x_{prep}$ [nmol/mol]	Prep. date	Analysis date	Sensitivity	Sensitivity Ratio*	$RD_{D155876}^{B,C,D,m}$ [%]	Matrix
D015249	344.38	2014.1.11	2014.3.14	0.8496	1	-	N <sub>2</sub>
D155876	344.02	2014.1.11	2014.3.14	0.8646	1.0176	0.0000	Air
D015286	353.26	2012.7.21	2014.3.14	0.8724	1.0268	0.9041	Air
D015224	356.33	2012.7.21	2014.3.14	0.8877	1.0448	2.6730	Air
D015230	351.52	2012.7.21	2014.3.14	0.8807	1.0356	1.7689	Air
D015275	342.43	2012.7.21	2014.3.14	0.8722	1.0270	0.9237	Air

11 \* Ratio of the sensitivities of corresponding raw to D015249 of CO/N<sub>2</sub> in first raw (Ratio<sub>air, nth</sub> = S<sub>air, nth</sub>/S<sub>D015249</sub>)

12

13

14 In 2014 fall meeting, It was agreed that KCRV of each cylinder be a preparation  
 15 value and associated uncertainty including stability drift, which (u<sub>stab.</sub>) is regarded to be an  
 16 interval (u<sub>stab.</sub>) between, before ( $x_{prep}$  in Table 4), and after ( $x_{2nd}$  in Table 4) shipping. Therefore  
 17 reference value is adopted as the preparation value and its total uncertainty includes changes  
 18 in CO mole fraction due to mixture drift of positive direction in 10 months. The KCRV and  
 19 associated expanded uncertainty of each cylinder are listed in Table 7.

20

1  
2

Table 7. KCRV and its Uncertainty budget including stability change

Laboratory	Cylinder	$x_{prep}$ [nmol/mol]	$U_{prep\_i}$ [nmol/mol]	$U_{stab.}$ ( $x_{2nd} - x_{prep}$ ) [nmol/mol]	$U_{prep\_f}$ [nmol/mol]
LNE	D015224	356.33	1.33	7.32	7.44
NIST	D015230	351.52	1.31	6.10	6.23
NOAA	D015283	350.81	1.31	3.47	3.71
FMI	D015285	348.86	1.30	2.96	3.23
NPL	D015215	349.97	1.30	3.02	3.29
JRC	D015217	350.83	1.31	3.13	3.39
NMIJ	D015275	342.43	1.27	3.04	3.30
KRISS	D015286	353.26	1.31	3.48	3.72
NIM	D015220	352.35	1.31	3.32	3.57
VNIM	D015223	351.35	1.31	3.07	3.34

3

#### 4 Measurement results

5 The measurement and calibration methods used by the participating laboratory in this comparison  
6 are listed in Table 8. Participants used laser spectroscopy, GC with various detectors or NDIR for  
7 the gas analysis. It is shown that laser absorption spectroscopy such as cavity ring down  
8 spectrometer (CRDS) was preferred by many NMIs. Judging from Figure 4, it was difficult to  
9 find which measurement technique would be the best for CO/air analysis. All participants except  
10 FMI (NPL standard) used their own standards.

11


12 Table 8. Summary of the measurement methods of the participants

Laboratory	Cylinder	Measurement period	Calibration standards	Instrument calibration	Measurement technique
LNE	D015224	Sep. 2012	Own standards	single point	QC-TILDAS
NIST	D015230	Sep. 2012	Own standards	multiple point	*CRDS
NOAA	D015283	Sep to Dec.	WMO-2004	Two point	**ICOS
FMI	D015285	Nov. 2012	NPL	Two point bracket	*CRDS
NPL	D015215	Dec. 2012 to Jan. 2013	Own standards	single point	*CRDS
JRC	D015217	Sep. 2012	Own standards	multiple point	***NDIR

			After 1 step dilution		
NMIJ	D015275	Sep. 2012	Own standards	multiple point	GC-HgO
KRISS	D015286	Jul. 2012 to Sep.	Own standards	single point	GC/FID/Methanat or
NIM	D015220	Mar. 2013	Own standards	single point	GC/FID/Methanat or
VNIIM	D015223	Sep. 2012	Own standards	single point	GC/FID/Methanat or

1 \*CRDS: Cavity ring-down spectrometer, \*\*ICOS: Off axis integrated cavity output spectroscopy, \*\*\*NDIR: Non dispersive  
2 infrared analysis method

3  
4 Preparation values and participants' reported values in this comparison are summarized in Table  
5 9 and shown in Figure 4.



6  
7 Figure 4. CO mole fractions between reported by participating laboratories (red circle) as well as KCRV (black  
8 square), where the vertical bars represent the expanded uncertainty reported by participants (red bar) and KCRV  
9 (black bar)

10  
11  
12

1 The difference between reported and prepared was listed in Table 9.

2 Table 9. Measurement Results of CCQM K84

Laboratory	Cylinder	$x_{\text{prep}}$	$u_{\text{prep\_tot}}$	$x_{\text{lab}}$	$U_{\text{lab}}$	$\Delta x$	$u(\Delta x)$
		[nmol/mol]	[nmol/mol]	[nmol/mol]	[nmol/mol]	$x_{\text{lab}} - x_{\text{prep}}$ [nmol/mol]	$u(x_{\text{lab}} - x_{\text{prep}})$ [nmol/mol]
LNE	D015224	356.33	3.72	356.6	1.1	0.3	3.8
NIST	D015230	351.52	3.12	358.6	2.0	7.1	3.3
NOAA	D015283	350.81	1.86	346.0	4.7	-4.8	3.0
FMI	D015285	348.86	1.62	351.84	5.52	2.98	3.20
NPL	D015215	349.97	1.65	355.4	7.0	5.4	3.9
JRC	D015217	350.83	1.7	351.09	13.10	0.26	6.77
NMIJ	D015275	342.43	1.65	341.26	2.46	-1.17	2.06
KRISS	D015286	353.26	1.86	353.25	1.06	-0.01	1.94
NIM	D015220	352.35	1.79	355.5	3	3	3
VNIIM	D015223	351.35	1.67	354	5	3	3

3  
4

## 5 Degrees of equivalence (DoE)

6 For the sake of consistency between the results of the participating laboratories and the  
7 KCRV, a degree of equivalence ( $d_i$ )<sup>(5)</sup> is expressed as

$$8 \quad d_i = x_i - x_{i,\text{KCRV}}, \quad (2).$$

9 In the above equation,  $x_{i,\text{KCRV}}$  identifies the key comparison reference value which is already  
10 mentioned as  $x_{\text{prep}}$ , and  $x_i$  is the result of laboratory  $i$ . Therefore the standard uncertainty of  $d_i$   
11 based on Table 9 can be expressed as:

$$12 \quad u^2(d_i) = u^2_{i,\text{lab}}(x_i) + u^2_{i,\text{prep\_tot}}(x_{i,\text{KCRV}}) \quad (3).$$


13

14 Assuming that the terms in equation (3) are uncorrelated, the degrees of equivalence  $d_i \pm$

15  $U(d_i)$  are presented in Figure 5, where the solid squares represent the  $d_i$  and the vertical bars

16 indicate the associated expanded uncertainty ( $k = 2$ ).





1

2 **Figure 5.** Differences between participants' results and the KCRV for the comparison, where the vertical bar  
 3 represents the expanded uncertainty,  $U(d_i)$ , at the 95 % level of confidence.  
 4

5

Table 10. Degrees of equivalence with uncertainties ( $k=2$ ) of CCQM K84

Laboratory	Cylinder	$d_i$ [nmol/mol]	$U(d_i)$ [nmol/mol]
LNE	D015224	0.3	7.5
NIST	D015230	7.1	6.5
NOAA	D015283	-4.8	6
FMI	D015285	2.98	6.4
NPL	D015215	5.4	7.7
JRC	D015217	0.26	13.53
NMIJ	D015275	-1.17	4.12
KRISS	D015286	-0.01	3.87
NIM	D015220	3	5
VNIIM	D015223	3	6

6

7

## 1 **Conclusions**

2 An increase in CO mole fraction was observed for all cylinders used for this comparison.  
3 Since drift rates were not uniform, it was decided to use the prepared values as the KCRV,  
4 and to include mole fraction changes in the KCRV uncertainty.

5 Most of the results in this key comparison are consistent with their KCRV within  
6 uncertainties. In order to achieve the WMO recommended comparability goal of  $\pm 2$   
7 nmol/mol, the stability of CO in air requires further attention.

8  
9 This key comparison supports the measurement capability of CO at 350 nmol/mol. The  
10 results of the comparison identify measurement equivalence between NMIs and WMO. This  
11 key comparison supports the measurement capability of CO in the range of 50 nmol/mol to  
12 1000 nmol/mol.

## 13 **Acknowledgements**

14 This work was funded by the Korea Meteorological Administration Research and  
15 Development Program under Grant No. KMIPA 2015-2032 and the Korea Research Institute  
16 of Standards and Science Program.

## 18 **References**

- 19 1. Novelli, P. C., J. W. Elkins and L. P. Steele, The Development and Evaluation of a  
20 Gravimetric Reference Scale For Measurements of Atmospheric Carbon Monoxide,  
21 Journal of Geophysical Research-Atmospheres, 96 (1991), D7, 13109-13121,  
22 JD01108.
- 23 2. ESRL Global Monitoring Division, Global view of CO,  
24 [http://www.esrl.noaa.gov/gmd/ccgg/globalview/co/co\\_intro.html](http://www.esrl.noaa.gov/gmd/ccgg/globalview/co/co_intro.html),
- 25 3. GAW report No. 213, 17th WMO/IAEA Meeting on Carbon Dioxide, Other  
26 Greenhouse Gases and Related Tracers Measurement Techniques (GGMT-2013),

1 World Meteorological Organization (WMO), Edited by Pieter Tans and Christoph  
2 Zellweger

3 4. International organization for standardization, ISO 6142.2, Gas analysis, “Preparation  
4 of calibration gas mixtures, Gravimetric method”, ISO, Third edition, 2013(E)

5 5. CCQM Guidance note: Estimation of a consensus KCRV and associated Degrees of  
6 Equivalence, Version 10 (2013-04-12).

7

8

9

1 Appendix: Measurement Reports

## 2 Report Form Carbon monoxide in synthetic air

3 Laboratory name: National Metrology Institute of Japan

4 Cylinder number: D015275

5

### 6 Measurement #1

Component	Date (dd/mm/yy)	Result ( $\mu\text{mol/mol}$ )	Standard deviation (% relative)	number of replicates
CO	03/09/12	339.85	0.33	9
CO	03/09/12	340.42	0.40	9
CO	04/09/12	341.93	0.32	9
CO	04/09/12	341.70	0.52	9
CO	05/09/12	343.48	0.53	9
CO	07/09/12	340.91	0.32	9
CO	07/09/12	340.54	0.36	9

7

8

9

### 10 Results

Component	Result ( $\mu\text{mol/mol}$ )	Expanded Uncertainty ( $\mu\text{mol/mol}$ )	Coverage factor <sup>1</sup>
CO	341.26	2.46	2

11

## 12 Method Description Forms

---

13

### 14 Details of the measurement method used:

### 15 Reference Method:

---

<sup>1</sup> The coverage factor shall be based on approximately 95% confidence.

1 Carbon monoxide concentration in synthetic air has been quantified using Peak performer 1 gas  
2 chromatography with reducing compound photometer (GC-HgO). Nitrogen purified by getter type  
3 purifier was used as carrier gas and the flow rate was 25 ml/min. Column temperature used is 100 °C.  
4 Sample was introduced to GC-HgO using 5mL sample loop and injection volume was corrected by  
5 sample temperature and pressure. The sample temperature is temperature of sample loop and the  
6 pressure is obtained from the downstream pressure gauge of sample loop.

## 8 **Calibration standards:**

### 9 Preparation method

10 Five calibration standards were used for the determination of carbon monoxide in synthetic air. The  
11 standards were prepared from pure carbon monoxide, pure nitrogen, and pure oxygen in accordance  
12 with ISO6142:2001 (Gas analysis-Preparation of calibration gases-Gravimetric method. Pure carbon  
13 monoxide was from Sumitomo Seika Chemicals Company Limited and pure nitrogen and oxygen  
14 from Japan Fine Products. Four-step dilution was used to make the mixtures, with carbon monoxide  
15 concentration of 23000-29000 μmol/mol, 610-670 μmol/mol, 13-16 μmol/mol, and 300-370  
16 vmol/mol. Oxygen was added in the last-step dilution. Table 1 shows gravimetric value and expanded  
17 uncertainty of the calibration standards.

18  
19 Table 1. Gravimetric value and expanded uncertainty in calibration standards

Cylinder number	Gravimetric Value (nmol/mol)	Expanded uncertainty [k=2] (nmol/mol)
CPC00873	310.55	0.32
CPB16345	371.07	0.32
CPC00420	300.99	0.32
CPC00414	348.78	0.32
CPC00418	328.87	0.32

### 21 Purity analysis

22 The impurities of carbon monoxide, nitrogen, and oxygen were determined by analytical methods and  
23 the amount of the major component is conventionally determined from the following equation,

$$24 \quad x_{pure} = 1 - \sum_{i=1}^N x_i$$

25 where

26  $x_i$ : the mole fraction of impurity  $i$ , determined by analysis;

27 N: the number of impurities likely to be present in the final mixture;

1  $x_{pure}$ : the mole fraction “purity” of the “pure” parent gas.

2 Table 2-4 shows the results of purity analyses. The purities of carbon monoxide, nitrogen and oxygen  
3 were not contributed to the uncertainty of the standard gases prepared by gravimetric method.

4

5 Table 2. Purity table of high-purity carbon monoxide used as parent gas

<i>Components</i>	<i>Applied concentration (<math>\mu\text{mol/mol}</math>)</i>	<i>Standard uncertainty (<math>\mu\text{mol/mol}</math>)</i>	<i>Distribution</i>	<i>Analytical method</i>
H <sub>2</sub>	2.87	0.61	Normal	Micro GC-TCD
He	19.46	0.24	Normal	Micro GC-TCD
N <sub>2</sub>	4.27	0.14	Normal	Micro GC-TCD
O <sub>2</sub>	1.65	0.19	Normal	Micro GC-TCD
CO <sub>2</sub>	0.39	0.22	Rectangular	Micro GC-TCD
N <sub>2</sub> O	0.42	0.24	Rectangular	Micro GC-TCD
CH <sub>4</sub>	0.75	0.44	Rectangular	Micro GC-TCD
H <sub>2</sub> O	0.33	0.02	Rectangular	Capacitance-type moisture meter
<b>CO</b>	999997.36			

6

7 Table 3. Purity table of high-purity nitrogen as parent gas

<i>Components</i>	<i>Applied concentration (<math>\mu\text{mol/mol}</math>)</i>	<i>Standard uncertainty (<math>\mu\text{mol/mol}</math>)</i>	<i>Distribution</i>	<i>Analytical method</i>
O <sub>2</sub>	0.819	0.473	Rectangular	Micro GC-TCD
CO <sub>2</sub>	0.010	0.006	Rectangular	FT-IR
CH <sub>4</sub>	0.026	0.015	Rectangular	FT-IR
CO	0.00029	0.00016	Rectangular	GC-HgO
H <sub>2</sub> O	0.439	0.253	Rectangular	Capacitance-type moisture meter
<b>N<sub>2</sub></b>	999998.48			

8

9

10 Table 4. Purity table of high-purity oxygen as parent gas

<i>Components</i>	<i>Applied concentration (<math>\mu\text{mol/mol}</math>)</i>	<i>Standard uncertainty (<math>\mu\text{mol/mol}</math>)</i>	<i>Distribution</i>	<i>Analytical method</i>
N <sub>2</sub>	0.090	0.052	Rectangular	GC-TCD
Ar	0.087	0.050	Rectangular	GC-TCD

CO <sub>2</sub>	0.056	0.004	Normal	FT-IR
CH <sub>4</sub>	0.003	0.002	Rectangular	FT-IR
CO	0.00179 (NSU 12-04) 0.00262 (NSN 75-82)	0.00016 (NSU 12-04) 0.00020 (NSN 75-82)	Normal	GC-HgO
H <sub>2</sub> O	0.439	0.253	Rectangular	Capacitance-type moisture meter
<b>O<sub>2</sub></b>	<b>999999.32</b>			

1

2 Concentrations of carbon monoxide in nitrogen and oxygen were also measured in detail by FTIR  
3 because concentration of carbon monoxide in nitrogen and oxygen, which is close to detection limit,  
4 contribute to uncertainty of gravimetric value. In FTIR with path length of 10m and MCT detector,  
5 measurement of carbon monoxide in nitrogen and oxygen was performed at high sample pressure of  
6 4000 hPa and integration of 16384 times to achieve high detection sensitivity. The concentration of  
7 carbon monoxide obtained using FTIR was agreed with that obtained using GC-HgO within  
8 uncertainty.

9

#### 10 **Instrument calibration:**

11 Instrument calibration is performed using NMIJ primary standard gas mixtures. A calibration curve  
12 was made between 300 and 370 nmol/mol by measuring the five calibration standards listed in table 1.  
13 NMIJ primary standards. The calibration curve was obtained using least square method.

14

#### 15 **Sample handling:**

16 The sample cylinder had been stood at room temperature for more than a week after arrival. Each  
17 cylinder was equipped with a stainless steel pressure regulator that was purged several times by  
18 sequential evacuation and pressurisation with the gas mixture used. Samples were transferred to  
19 sample loop at flow rate of 80 ml/min using mass-flow controller. Effect of sample pressure and  
20 temperature were corrected. The pressure was measured by pressure sensor at downstream of sample  
21 loop and the temperature was ambient temperature near sample loop.

22

23

#### 24 **Uncertainty:**

25 The uncertainty used for the calibration mixtures contains all sources of gravimetric preparation.  
26 Uncertainty for stability is not included because no instability has been detected. The uncertainty in  
27 measurement is calculated based on repeatability of the acquired area.

28

#### 29 **Detailed uncertainty budget:**

1 Please include a list of the uncertainty contributions, the estimate of the standard uncertainty,  
 2 probability distributions, sensitivity coefficients, etc.

3

4 Typical evaluation of the measurement uncertainty for CO:

Quantity $X_i$	Estimate $x_i$	Evaluation Type (A or B)	Distribution	Standard uncertainty $u(x_i)$	Sensitivity coefficient $c_i$	Contribution $u_i(y)$
References		A	Gaussian	0.16 $\mu\text{mol/mol}$		
Repeatability		A	Gaussian	1.22 $\mu\text{mol/mol}$		
Combined standard uncertainty				1.23 $\mu\text{mol/mol}$		

5

6



# Report Form Carbon monoxide in synthetic air

---

Laboratory name: NIST

Cylinder number: D015230

Analyst: Dr. Gerald Mitchell

The Gas Metrology Group of the Analytical Chemistry Division at NIST participated in the CCQM K84 Key Comparison. One (1) compressed gas cylinder containing carbon monoxide (CO) in a balance of air (cylinder number D015230) was analyzed as part of the comparison sponsored by the Gas Analysis Working Group of the Consultative Committee for Amount of Substance (CCQM). The mixture was prepared gravimetrically according to the CCQM-K84 protocol by the Korea Research Institute of Standard and Science (KRISS).

## Analytical Methods

**Carbon Monoxide Analysis:** The carbon monoxide was analyzed using a Los Gatos Model N<sub>2</sub>O/CO-23d analyzer (NIST # 642093). A computer operated gas analysis system (COGAS # 14) was used to deliver the sample stream to the analyzer. Prior to beginning, each analysis the sample line and regulator of each cylinder was purged five (5) times. The analyzer was used to measure the response ratio of each primary standard gas mixture (PSM) cylinder to that of control cylinder (FF10258). During each analytical run, the sample has a purge time of 1.5 minutes before data collection. The analyzer's internal pump used to draw the sample into the analysis cell. The CCQM cylinder and the PSMs listed below were measured against the control cylinder six times during two different analytical periods.

## Calibration Standards

Six NIST gravimetrically prepared primary reference materials ranging in concentration from 0.200  $\mu\text{mol/mol}$  to 0.500  $\mu\text{mol/mol}$  were used in this analysis. The PSMs are listed below:

<u>Cylinder Number</u>	<u>Concentration (<math>\mu\text{mol/mol}</math>)</u>	<u>Uncertainty (<math>\mu\text{mol/mol}</math> k=2)</u>
FF10204	0.2015	0.0019
FF10217	0.2526	0.0019
FF10221	0.2996	0.0019
FF10235	0.3483	0.0019
FF10226	0.3940	0.0018

1                    FF10228                    0.5032                    0.0018  
2

3    These standards were prepared from three different parent mixtures but all with the same source of  
4    balance gas (air).    The table 1 gives an assay of the air cylinders used to prepare the standards.

5    \_\_\_\_\_

1 **Table 1.** Assay of Pure Cylinders

Purity Pure Air (CC101252)	Concentration ( $\mu\text{mol/mol}$ )	Uncertainty ( $\mu\text{mol/mol}$ )	Purity Pure Air (CC2794)	Concentration ( $\mu\text{mol/mol}$ )	Uncertainty ( $\mu\text{mol/mol}$ )
N2 (difference)	781000	5	N2 (difference)	780930	9
CO (measured)	0.0077	0.0020	CO (measured)	0.0081	0.0020
CO2 (measured)	397.97	0.075	CO2 (measured)	393.2	0.051
CH4 (measured)	0.0017	0.0009	CH4 (measured)	0.0017	0.0009
Ar (measured)	9352.0	3.0	Ar (measured)	9352.0	8.0
O2 (measured)	209250	3.4	O2 (measured)	209325	3.4
Purity Pure Air (CC73041)	Concentration ( $\mu\text{mol/mol}$ )	Uncertainty ( $\mu\text{mol/mol}$ )	Purity Pure Air (CC81188)	Concentration ( $\mu\text{mol/mol}$ )	Uncertainty ( $\mu\text{mol/mol}$ )
N2 (difference)	780971	9	N2 (difference)	780885	10
CO (measured)	0.0081	0.0020	CO (measured)	0.0064	0.0020
CO2 (measured)	393.15	0.05	CO2 (measured)	397.10	0.50
CH4 (measured)	0.0017	0.0009	CH4 (measured)	0.0017	0.0009
Ar (measured)	9352.0	8.0	Ar (measured)	9349.0	6.0
O2 (measured)	209284	3.8	O2 (measured)	209369.0	8.1

2

3 **Instrument Calibration:** The instrument used in this study is a Los Gatos Model N<sub>2</sub>O/CO-23d  
 4 analyzer. It was calibrated using six gravimetrically prepared PSMs ranging in concentration from  
 5 0.200  $\mu\text{mol/mol}$  to 0.500  $\mu\text{mol/mol}$ . The CCQM sample (D015230) was included in the analysis  
 6 with the PSMs. They were all compared to a control a minimum of six times during each of the two  
 7 analytical periods. The analytical scheme used was, Control – PSM Standard (1) – Control – PSM  
 8 Standard (2) Control, etc. The procedure called for each cylinder to have a 1.5 minutes period of  
 9 equilibration and two-minute data collection period. A calibration curve using all data generated  
 10 during the two analytical periods was generated using PSM Validator (table 2).

11 **Sample Handling**

12 This analysis is to quantify the CO in a single CCQM-K84 cylinder (D015230). The sample was fitted  
 13 with a regulator adapted to the fitting supplied by KRIS. The measurements were automated  
 14 using NIST data system (# 631251) and a computer operated gas analysis system (COGAS # 14).  
 15 Prior to starting each set of analyses the regulator was flushed five times. The output pressure of  
 16 each regulator was set so that the slip stream feeding the analyzer has an excess of 200 ml/min.

17 **Amount-of-Substance Fraction**

18 The amount-of-substance fraction (concentration) for carbon monoxide in the K-84 PRM cylinder  
 19 number D015230 is given in table 3. All measured data and calculations for this CCQM key  
 20 concentration have been reviewed for sources of systematic and random errors. The uncertainty  
 21 of the concentration is expressed as an expanded uncertainty,  $U = k_c$  with a coverage factor  $k$  equal  
 22 to 2. The true concentration is asserted to lie within the interval defined  $(0.3586 \pm 0.0020)$   
 23  $\mu\text{mol/mol}$  with a level of confidence of approximately 95 % [1]. The uncertainty has been  
 24 expanded to 0.0020  $\mu\text{mol/mol}$ , as this corresponds to the uncertainty of our primary standards.

1 \_\_\_\_\_

1 **Table 2.** CCQM-K84 Cylinder D015230 Validation Data

GENLINE - Linear ( $y=b_0+b_1*x$ )					
	Value	Std Error			
b0	0.009141356	0.00265624			
b1	0.296060862	0.002296321			
cov(b0,b1)		-5.8386E-06			
rms residual error		1.217144142			
X	Y	X-Solution	Y-Solution	uTest	Cylinder ID
0.6681	0.2051	0.6681	0.2069	PASS	FF10204
0.8176	0.2526	0.8176	0.2512	PASS	FF10217
0.9708	0.2996	0.9709	0.2966	PASS	FF10221
1.1503	0.3483	1.1503	0.3497	PASS	FF10235
1.3069	0.3940	1.3069	0.3961	PASS	FF10226
1.6662	0.5032	1.6662	0.5024	PASS	FF10228
Xin	uXin	Yeval	uYeval	Grav Value (k=2)	Cylinder ID
1.0000	0.0000	0.30520	0.00081	0.3042 ± 0.0039	FF10258
1.1804	0.0003	0.35861	0.00079	Unknown	D015230

2

3

4 **Table 3.** Amount-of-Substance Fraction (Concentration) for CCQM-K-84 Cylinder # D015230

5	Cylinder	Measured Value
6	<u>Contents</u>	<u>(September 2012)</u>
7	Carbon monoxide in Air	(0.3586 ± 0.0020) µmol/mol

8

9

10

1 Appendix A.

2

# 3 Report Form Carbon monoxide in synthetic air

4

5

6 Laboratory name: LNE

7 Cylinder number: D0152249

8

9

10 Measurement #1

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	21/09/2012	356.83	0.03	3

11

12 Measurement #2

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	27/09/2012	356.82	0.04	3

13

14 Measurement #3

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	28/09/2012	356.33	0.07	3

15

16 Measurement #4

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	01/10/2012	357.07	0.04	3

17

18 Measurement #5

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	03/10/2012	356.02	0.05	3

19

1 Measurement #6

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	04/10/2012	357.10	0.10	3

2

3

4

5

6 Results

7

Component	Result (nmol/mol)	Expanded Uncertainty (nmol/mol)	Coverage factor <sup>2</sup>
CO	356.6	1.1	2

8

9

10

11 **Method Description Forms**

12

13 Please complete the following data regarding the description of methods and the uncertainty  
14 evaluation.

15

16 Details of the measurement method used:

17

18 **Reference Method:**

19

20 A Quantum Cascade Tunable Infrared Laser Differential Absorption Spectrometer (QC-  
21 TILDAS) developed by Aerodyne Research Inc. is used to quantify CO concentration.  
22 This analytical system consists of four pulsed lasers (one for each analyte), an optical system  
23 with a 210 m path length absorption cell (Herriott type) and a cooled HgCdTe (MCT) detect  
24 or.

25

26 **Calibration standards:**

27

---

<sup>2</sup> The coverage factor shall be based on approximately 95 % confidence.

1 Gravimetric gas mixtures were prepared in accordance with the international standard EN  
2 ISO 6142.

3 Two gas mixtures (CO/Air 0008 and CO/Air 0009) at about 350 nmol/mol of carbon  
4 monoxide in synthetic air were prepared using gravimetric method.

5 Each preparation needed 4 steps (4 gravimetric gas mixtures) to obtain the nominal  
6 concentration of 350 nmol/mol.

7 The mass of each component has been measured by comparison between the mass of the  
8 cylinder and a standard cylinder (tare) with a comparator METTLER AX3200 with a  
9 resolution of 0.1 mg and standard masses.

### 11 Instrument calibration:

12  
13 The background of the QC-Laser spectrometer was made with zero air. Then, the  
14 spectrometer responses were recorded for gravimetric gas mixture and for the unknown gas  
15 mixture. The amount fraction of the unknown gas mixture is determined using the amount  
16 fraction of the gravimetric gas mixture and the ratio of the spectrometer responses to the  
17 gravimetric gas mixture and the unknown gas mixture.

18 This sequence is repeated three times during a day and on 2 days or 4 days respectively with  
19 the CO/Air 0008 gravimetric gas mixture and the CO/Air 0009 gravimetric gas mixture.

### 22 Sampling handling:

23  
24 Cylinders were maintained at a laboratory temperature of  $(21 \pm 2)$  °C throughout the period  
25 of analysis.

26 Samples were introduced into the analyzer at atmospheric pressure (excess flow was passed  
27 to vent) using a low volume gas regulator.

### 29 Uncertainty:

#### 31 **1) Gravimetric gas mixtures uncertainties :**

32 ✓

33 ✓ As explained before the preparation of the 2 gravimetric gas mixtures at about 350  
34 nmol/mol needed the preparation of 4 gravimetric gas mixtures.

35 The last 2 gravimetric gas mixtures at about 350 nmol/mol were obtained by diluting  
36 gravimetrically the third gravimetric gas mixture (CO at 10 μmol/mol – CO/N2 0061) with  
37 argon, oxygen and nitrogen.

38  
39 ✓ *Gravimetric gas mixture n° CO/Air 0008*



1  
2  
3

- ✓  
• **Purity tables of each component**

✓ Component	Concentration (mol/mol)	Uncertainty (mol/mol)
CO	0.00000998813483	0.00000000669398
N <sub>2</sub>	0.99998759588702	0.00000007209187
Ar	0.00000235021068	0.00000006884169
methane	0.00000002549589	0.00000001412405
H <sub>2</sub>	0.00000002549090	0.00000001412405
CO <sub>2</sub>	0.00000000696628	0.00000000098025
O <sub>2</sub>	0.00000000511464	0.00000000284441
H <sub>2</sub> O	0.00000000201496	0.00000000117632

4  
5

Purity table of CO/N2 0061

✓ Component	Concentration (mol/mol)	Uncertainty (mol/mol)
Ar	0.9999998430	0.000000087
N <sub>2</sub>	0.000000150	0.000000087
O <sub>2</sub>	0.000000005	0.0000000029
methane	0.0000000015	0.000000001
CO <sub>2</sub>	0.0000000005	0.000000001
CO	0.000000000	.0000000005
H <sub>2</sub> O	0.000000000	0.000000001

6  
7  
8  
9  
10  
11  
12

Purity table of Argon (BIP+ n°A106273)

Component	Concentration (mol/mol)	Uncertainty (mol/mol)
O <sub>2</sub>	0.999997842	0.000001155
CO	0.00000000283	0.0000000008
CO <sub>2</sub>	0.000000139	0.000000005
N <sub>2</sub>	0.000002000	0.0000011547
methane	0.0000000015	0.000000001
NO <sub>2</sub>	0.000000015	0.0000000087
H <sub>2</sub> O	0.000000000	0.000000001

13  
14

Purity table of Oxygen (Alphagaz n°20026895)

Component	Concentration (mol/mol)	Uncertainty (mol/mol)
N <sub>2</sub>	0.999999711	0.000000037

O <sub>2</sub>	0.000000005	0.0000000029
Ar	0.0000002566	0.000000034
CO <sub>2</sub>	0.0000000005	0.000000001
methane	0.0000000015	0.000000001
H <sub>2</sub>	0.000000025	0.0000000144
H <sub>2</sub> O	0.0000000002	0.0000000012
CO	0.000000000	0.0000000005

Purity table of Nitrogen (BIP+ n°293523)

• **Mass of each component for preparing the gravimetric gas mixture CO/Air 0008**

Component	Mass (g)	Uncertainty (g)
CO/N2 0061	44.888	0.013
Argon BIP + n°A106273	17.952	0.013
Oxygen alphagaz n°20026895	306.523	0.013
Nitrogen BIP+ n°293523	941.596	0.015

✓

• **Composition of the gravimetric gas mixture CO/Air 0008**

Component	Concentration (mol/mol)	Uncertainty (mol/mol)
Carbon monoxide	$354.35 \cdot 10^{-9}$	$0.49 \cdot 10^{-9}$
Argon	$0.99331 \cdot 10^{-2}$	$0.00069 \cdot 10^{-2}$
Oxygen	$21.1726 \cdot 10^{-2}$	$0.00079 \cdot 10^{-2}$
Nitrogen and others impurities	balance	-

1 ✓ Gravimetric gas mixture n° CO/Air 0009

2 ✓

3 • Purity tables of each component

4

✓ Component	Concentration (mol/mol)	Uncertainty (mol/mol)
CO	0.00000998813483	0.00000000669398
N <sub>2</sub>	0.99998759588702	0.00000007209187
Ar	0.00000235021068	0.00000006884169
methane	0.00000002549589	0.00000001412405
H <sub>2</sub>	0.00000002549090	0.00000001412405
CO <sub>2</sub>	0.00000000696628	0.0000000098025
O <sub>2</sub>	0.00000000511464	0.00000000284441
H <sub>2</sub> O	0.00000000201496	0.00000000117632

5

Purity table of CO/N<sub>2</sub> 0061

6

✓ Component	Concentration (mol/mol)	Uncertainty (mol/mol)
Ar	0.9999998430	0.000000087
N <sub>2</sub>	0.000000150	0.000000087
O <sub>2</sub>	0.000000005	0.0000000029
methane	0.0000000015	0.000000001
CO <sub>2</sub>	0.0000000005	0.000000001
CO	0.000000000	.0000000005
H <sub>2</sub> O	0.000000000	0.000000001

7

Purity table of Argon (BIP+ n°A106273)

8

Component	Concentration (mol/mol)	Uncertainty (mol/mol)
O <sub>2</sub>	0.999997842	0.000001155
CO	0.00000000283	0.0000000008
CO <sub>2</sub>	0.000000139	0.000000005
N <sub>2</sub>	0.000002000	0.0000011547
methane	0.0000000015	0.000000001
NO <sub>2</sub>	0.000000015	0.0000000087
H <sub>2</sub> O	0.000000000	0.000000001

9

Purity table of Oxygen (Alphagaz n°20026895)

10

Component	Concentration (mol/mol)	Uncertainty (mol/mol)
N <sub>2</sub>	0.999997648	0.000000073
O <sub>2</sub>	0.000000005	0.0000000029
Ar	0.000002288	0.000000070
CO <sub>2</sub>	0.000000007	0.000000001
methane	0.000000025	0.0000000144

H <sub>2</sub>	0.000000025	0.0000000144
H <sub>2</sub> O	0.000000002	0.0000000012
CO	0.000000000	0.0000000005

Purity table of Nitrogen (BIP+ n°082420)

- **Mass of each component for preparing the gravimetric gas mixture CO/Air 0009**

Component	Mass (g)	Uncertainty (g)
CO/N2 0061	44.606	0.013
Argon BIP + n°A106273	17.764	0.013
Oxygen alphagaz n°20026895	300.797	0.013
Nitrogen BIP+ n°082420	939.722	0.015

- **Composition of the gravimetric gas mixture CO/Air 0009**

Component	Concentration (mol/mol)	Uncertainty (mol/mol)
Carbon monoxide	$354.15 \cdot 10^{-9}$	$0.49 \cdot 10^{-9}$
Argon	$0.98874 \cdot 10^{-2}$	$0.00069 \cdot 10^{-2}$
Oxygen	$20.8975 \cdot 10^{-2}$	$0.00080 \cdot 10^{-2}$
Nitrogen and others impurities	balance	-

**2) Detailed uncertainty budget:**

Uncertainty source	Estimate $x_i$ (nmol/mol)	Assumed distribution	Standard uncertainty $u(x_i)$ (nmol/mol)
Mean concentration obtained by comparison with the gravimetric gas mixture CO/Air 0008	356.18	Standard deviation of the values	0.25
Mean concentration obtained by comparison with the gravimetric gas mixture CO/Air 0009	356.95	Standard deviation of the values	0.22
Gravimetric gas mixture concentration (CO/Air 0008)	354.35	-	0.49
Gravimetric gas mixture concentration (CO/Air 0009)	354.15	-	0.49

The concentration of the unknown gas mixture D0152249 is the mean concentration of the 2 mean concentrations obtained by comparison with the 2 gravimetric gas mixtures CO/Air 0008 and CO/Air 0009:

1 
$$C_{D0152249} = \frac{356.18 + 356.95}{2} = 356.6 \text{ nmol / mol}$$

2

3 The uncertainty on the unknown gas mixture concentration is given by:

4 
$$u^2(C_{D0152249}) = \frac{1}{4} (u^2(C_{CO/Air\ 0008}) + s_{CO/Air\ 0008}^2 + u^2(C_{CO/Air\ 0009}) + s_{CO/Air\ 0009}^2 + 2 \times r \times u^2(C_{CO/Air\ 0008}) \times u^2(C_{CO/Air\ 0009}))$$

5 
$$u^2(C_{D0152249}) = \frac{1}{4} (0.49^2 + 0.25^2 + 0.49^2 + 0.22^2 + 2 \times 1 \times 0.485 \times 0.485) = 0.26$$

6 (r is equal to 1 because the 2 gravimetric gas mixtures are correlated)

7 And 
$$U(C_{D0152249}) = 1.1 \text{ nmol / mol}$$

8

# Report from Carbon monoxide in synthetic air

Laboratory name: Finnish Meteorological Institute

Cylinder number:

## Measurement #1

Component	Date (dd/mm/yy)	Results (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	07/11/2012	351,42	0,04 %	10

## Measurement #2

Component	Date (dd/mm/yy)	Results (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	08/11/2012	351,92	0,03 %	10

## Measurement #3

Component	Date (dd/mm/yy)	Results (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	08/11/2012	352,17	0,05 %	10

## Results

Component	Results (nmol/mol)	Expanded Uncertainty	Coverage factor
CO	351,84	1,57 %	2

1

2

3 Reference Method:

4 **Describe your instrument:**

5 The analyzer used for the measurements was N<sub>2</sub>O-CO-H<sub>2</sub>O analyzer by Los Gatos. It is  
6 based on cavity enhanced laser absorption technique.

7

8 **Calibration standards:**

9 The calibration standard was purchased from NPL, UK. The concentration of the gas  
10 standard was 20.09 µmol/mol in synthetic air with the standard uncertainty of 0.25 %.

11 **Instrument calibration:**

12 The calibration of the N<sub>2</sub>O-CO-H<sub>2</sub>O analyzer was made by dilution of the gas standard at the  
13 concentration level of CCQM-K84 nominal value. The calibration concentration of 326,4  
14 µmol/mol and 366,1 µmol/mol was used bracketing the CCQM-K84 concentration. The  
15 synthetic air was used for dilution of the gas standard. The purity of the synthetic air was  
16 examined by comparison of the respond of the instrument for synthetic air and to pure  
17 nitrogen gas. The linear regression line was fitted into the measured mean values of the  
18 analyzer against the value of the calibration concentration. The gas dilutor, Environics 6100,  
19 was calibrated against the flow measurement system of Molbloc laminar flow element by  
20 DHI, which was calibrated against the primary flow measurement system by MIKES.  
21 Linearity of the analyzer as well as the standard error of the estimate of the regression line  
22 was used to examine the performance of the analyzer during calibration. The lack of linearity  
23 was included into the uncertainty budget for the measurement results.

24 During the measurements the temperature and the pressure of the laboratory room was  
25 recorded. The changes of both quantities were within the acceptable range and no correction  
26 on the results based on these was made.

27

28 **Sampling handling:**

1 The CCQM-84 cylinder arrived in August when the ambient temperature was well above the  
 2 condensation temperature in the cylinder i.e. no condensation was expected for gas  
 3 compound in the cylinder. While in sending the cylinder back to KRISS a week of delay was  
 4 made in order to wait that the ambient temperature was decreased from – 20 °C to 0 °C. The  
 5 measurements were made between 7 to 8, November, 2012.

6 The pressure regulator (two stages) was connected into the cylinder. The pressure regulator  
 7 was flushed several time with the cylinder gas and finally it was conditioned both the primary  
 8 and secondary part of the pressure regulator overnight. The sample was injected directly to  
 9 the analyzer through the tube of stainless steel at room pressure. The excess of about 0.3 to  
 10 0.5 l/min was used to maintain the gas at room pressure.

11

12 **Uncertainty:**

13 Sources of uncertainty of the performance characteristics of the analyzer, gas dilutor, zero air  
 14 and the gas standard are listed in Table 5

15 Table 5 Uncertainty budget for the measurement results of CCQM-K84

16

Quantity Xi	Estimate Xi(%; nmol/mol)	Evaluation Type (A or B)	Distribution	Standard Uncertainty, u(xi)	Sensitivity coefficient, ci	Contribution ui(y)
Detection limit	0,40	B	Gaussian	0,40	1,00	0,40
Repeatability	0,9	A	Rectangular	0,52	1,00	0,52
Linearity	0,25 %	A	Rectangular	0,14 %	1	0,14 %
zero gas, purity	0,5	A	Rectangular	0,29	1,00	0,29
Gas standard	0,25 %	B	Gaussian	0,25 %	1,00	0,25 %
Gas dilution	0,70 %	B	Gaussian	0,70 %	1,00	0,70 %
Combined std uncertainty uc(xi)						0,78 %
Expanded uncertainty U(%)						1,57 %
Coverage factor, k						2

17

18

19

20



# VNIIM Report CCQM-K84: Carbon monoxide in synthetic air

Authors: L.A. Konopelko, Y.A. Kustikov, A.V. Kolobova, V.V. Pankratov, I.I. Wasserman,  
S.V. Zav'yalov, O.V. Efremova, M.V. Pavlov.

Laboratory: VNIIM, Research Department for the State Measurement Standards in the field  
of Physico-Chemical Measurements.

Cylinder number: D015223

## Measurement #1

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	06/09/2012	353.4	0.43	6

## Measurement #2

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	06/09/2012	353.3	0.40	6

## Measurement #3

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	12/09/2012	353.2	0.45	6

## Measurement #4

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	12/09/2012	353.3	0.69	6

## Measurement #5

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	20/09/2012	354.9	0.70	6

## Measurement #6

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	20/09/2012	353.9	0.76	6

1

2 **Result**

<b>Component</b>	<b>Result (nmol/mol)</b>	<b>Expanded Uncertainty (nmol/mol)</b>	<b>Coverage factor</b>
<b>CO</b>	<b>354</b>	<b>5</b>	<b>2</b>

3 Details of the measurement method used:

4

5 **Reference Method:**

6

7 Gas chromatography with FID and methanizer;

8 Instrument: Gas Chromatograph «Agilent 6890»;

9 Column: CaA 3 m x 3mm;

10 Carrier gas: helium 30 ml/min;

11 Oven conditions: 75 °C for 9 min;

12 Sample loop: 5 ml;

13 Data collection: by “ChemStation A.10.02” software.

14

15 **Calibration Standards:**

16

17 Calibration was performed using Primary Standard Gas Mixtures, prepared by the  
18 gravimetric method from pure substances, according to ISO 6142:2001 “Gas analysis -  
19 Preparation of calibration gas mixtures - Gravimetric method”.

20 Characteristics of pure substances used for preparation of the calibration gas mixtures are  
21 shown in the table 1.

22 Table 1 – Description of pure substances

Substance	Mole fraction (µmol/mol)	Standard uncertainty (µmol/mol)
Carbon monoxide	999861	22
Nitrogen	999999.0	0.4
Oxygen	999998.4	0.3
Argon	999999.50	0.10

23

24 Preparation from pure substances was carried out in 4 stages. On the first stage 3 CO/N<sub>2</sub> gas  
25 mixtures were prepared on the concentration level of 2,5 %. On the second and third stages  
26 these mixtures were diluted respectively to the concentration levels of 500 and 9 µmol/mol.  
27 Then the last mixtures were diluted to target concentration level of 0.3 µmol/mol.

1 The concentrations and standard uncertainties of carbon monoxide in these mixtures are  
2 shown below.

3

Cylinder number	Component	Concentration (nmol/mol)	Standard uncertainty due to weighing and purity (nmol/mol)
D249246	CO	300.0	1.0
D249373	CO	301.2	1.0
D249403	CO	300.8	1.0

4

5 All standard gas mixtures were prepared in aluminum cylinders (Luxfer) with Aculife IV +  
6 Aculife III treatment.

7

### 8 Instrument Calibration:

9

10 Single point calibration method was used to determine carbon monoxide concentration in the  
11 gas mixture to be investigated. Linearity of the carbon monoxide calibration curve  
12 (mathematical model of calibration curve  $y=kx$ ) in the range 0.300 – 0.500  $\mu\text{mol/mol}$  was  
13 checked preliminarily using Primary Standard Gas Mixtures.

14 Measurement sequence was in the order: standard-sample-standard-sample-standard (etc.).  
15 Temperature and pressure were not corrected during the calibration procedure due to cited  
16 above measurement sequence.

17

### 18 Sample Handling:

19

20 Prior to measurements cylinders were stabilized to room temperature. Each cylinder was  
21 equipped with a pressure regulator that had been adequately purged before the sample was  
22 transferred to the sample loop. The additional valve was installed between the outlet of the  
23 column and methanator order to remove oxygen before the methanator.

24

### 25 Uncertainty:

26

27 a) Uncertainty related to calibration standards, which takes into consideration uncertainty of  
28 the balance, weights and purity analysis of the parent gases, was calculated with the program  
29 developed in VNIIM on the base of ISO 6142:2001 “Gas analysis - Preparation of calibration  
30 gas mixtures - Gravimetric method”;

31 b) Uncertainty related to the measurements -  $S_x$  was calculated according to the formula

$$1 \quad S_x = \sqrt{S_R^2 + S_r^2 \frac{n-1}{n}}$$

$$2 \quad S_R = \sqrt{\frac{\sum_{i=1}^m (x_i - \bar{x})^2}{m(m-1)}}$$

$$3 \quad S_r = \sqrt{\frac{\sum_{i=1}^m S_{r_i}^2}{m}}$$

$$4 \quad S_{r_i} = \sqrt{\frac{\sum_{j=1}^n (x_{ij} - x_i)^2}{n(n-1)}}$$

5 where  $n$  – number of the results in one measurement series;

6  $m$  – number of the measurement serieses;

7  $x_{ij}$  – one of the single results in one of the measurement serieses;

8  $x_i$  – the average result in one measurement series;

9  $\bar{x}$  – overall mean.

10

11 Detailed uncertainty budget:

Uncertainty source $X_i$	Estimate $x_i$	Evaluation type (A or B)	Distribution	Standard uncertainty $u(x_i)$	Sensitivity coefficient $c_i$	Contribution $u_i(y)$ , %
Calibration standards (weighing + purity)	300.0 (nmol/mol)	A, B	Normal	1.0 (nmol/mol)	1.179	0.333
Between days and within day measurements	353.6 (nmol/mol)	A	Normal	1.92 (nmol/mol)	1	0.544
Combined standard uncertainty						0.638
Expanded uncertainty $k=2$						1.3

12

13

1

**JRC –ERLAP – CCQM-K84**

2

**Analysis of Carbon monoxide in Synthetic air at ambient level**

3

4

Cylinder code: KC10

5

Cylinder number: D015217

## Report Form Carbon monoxide in synthetic air

Laboratory name: JOINT RESEARCH CENTERCylinder number: KC10  
D015217

## Measurement #1

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	12/09/12	350,14	1,04	5

## Measurement #2

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	14/09/12	348,78	1,14	5

## Measurement #3

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	17/09/12	354,34	1,03	5

## Measurement #4

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	/			

## Measurement #5

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	/			

## Results

Component	Result (nmol/mol)	Expanded (nmol/mol) Uncertainty	Coverage factor <sup>1</sup>
CO	351,09	13,10	2

1 The coverage factor shall be based on approximately 95 % confidence.

## 2 Reference Method:

3 Analyzer: CO analyzer Horiba APMA 370

4 Method: Non dispersive infrared analysis method

5 Configuration: Stand alone

1 Data collection: Data acquisition system with homemade software. Sampling rate is around 2 sec  
2 and used to create 1 min averages which are stored in a file. Five consecutive 1 min averages build  
3 one measurement.

#### 4 **Calibration Standard:**

5 A dynamic dilution system with thermal mass flow controllers is used to produce gas-mixtures of a  
6 desired concentration. A Primary Reference Material (PRM) of  $300 \cdot 10^{-6}$  mol/mol, manufactured by  
7 VLS (NL), is diluted with homemade Zero Air. The PRM is produced according to ISO 6142:2001  
8 (Gravimetric preparation) with an expanded uncertainty of 0.3%. The Zero Air is generated from  
9 ambient air using a drying system, UV-lamp for oxidation, activated charcoal, molsiev, silicagel and a  
10 palladium converter of 380 deg. C for purification. It is compared to a Primary Reference Gas  
11 Mixture manufactured by NPL (UK) with an amount fraction of Carbon Monoxide  $\leq 15$  nmol/mol and  
12 found to be lower by around 10 nmol/mol.

#### 13 **Instrument calibration:**

14 The analyzer is calibrated right before analysis of a client cylinder. Zero air and three gas-mixtures of  
15 different concentrations are generated with the dynamic dilution system and measured for 15 min  
16 with the analyzer. The last 5 min are used to calculate the measurement result. Immediately after  
17 each single gas-mixture the flow-rate of the PRM and the Zero Air are measured independently with  
18 a Molbloc/Molbox1 system. The latter one was certified against gravimetric standards at LNE (F) in  
19 2011. Right after the calibration the client cylinder is measured for 15 min, the last five 1 min  
20 averages used to calculate the measurement result. As an experiment lasts for less than 2 hours,  
21 temperature/pressure corrections are not taken into account.

#### 22 **Sampling handling:**

23 Client cylinder was stabilized to ambient laboratory conditions over more than 2 weeks. A two-stage  
24 pressure regulator, manufactured by Veriflow, type 735, with a stainless steel needle valve was  
25 attached to the cylinder and evacuated before filled with the cylinder content. Teflon tubing with a  
26 t-piece was used to connect it to the analyzer allowing sampling at ambient conditions. In total,  
27 three independent calibrations and measurements were performed on Sept. 12<sup>th</sup>, 14<sup>th</sup> and 17<sup>th</sup>,  
28 2012, respectively.

#### 29 **Uncertainty:**

30 Calibration uncertainty:

31 The uncertainty of the PRM (expanded uncertainty 0.3%), of the Zero Air (7.5 nmol/mol), of the  
32 flow-rate (expanded uncertainty 0.3 % or 0.4 % of reading, depending on the flow-rate and used  
33 Molbloc) and the repeatability of the analyzer during the sampling of a calibration gas-mixture were

1 taken into account. Uncertainties on compression factors, used to calculate between mol/mol and  
2 ppb, were applied but found to be negligible. Analyzer resolution was neglected during calibration as  
3 calibration points in the range of 1000 to 3000 nmol/mol were used. Calibration uncertainties were  
4 combined using GUM – workbench software to obtain a target value and its associated uncertainty.  
5 An example of such a budget is attached. The target values, their uncertainties, the Zero – Air  
6 reading and all three calibration readings are entered into "bleast software" (based on ISO 6143) to  
7 calculate the calibration parameters: Linear regression with slope, intercept and associated  
8 uncertainties.

9           Analysis uncertainty:

10 GUM workbench was used again to calculate the final measurement result out of the calibration  
11 parameters, repeatability and analyzer resolution. The latter had to be included at the nominal  
12 concentration of the client cylinder, as analyzer resolution is 10 nmol/mol, hence it could have a  
13 significant influence on the final result. The three single measurement results and their uncertainty  
14 budgets are attached. Finally an average of the three measurements and the three expanded  
15 uncertainties was calculated to obtain the final result. The possibility to reduce the variable part of  
16 the uncertainty contributions due to the averaging of the final result was discarded.

17

18



1 National Institute of Metrology (NIM), China

2 **Transmission of International Comparison Results**

3

4

5

6 **The title of international comparison:** Carbon monoxide in synthetic air at ambient level

7

8 **Serial number for international comparison:** CCQM-K84

9 **Comparison experiment period:** January, 2013~April, 2013

10 **Experiment reporter:** HAN Qiao and HU Shuguo

11

12

13 **Phone number of experiment reporter:** +86-10-84252300

14 **E-mail:** hanqiao@nim.ac.cn, hushg@nim.ac.cn

15 **NIM address:** No.18, Bei-San-Huan Dong Str., Beijing 100013, China

16 **Phone/Fax number of Department of Metrology Services:** +86-10-64213104

17 **Phone/Fax number of Department of R&D and Planning (International Cooperation):**

18 +86-10-64218565

19 **E-mail:** yw@nim.ac.cn

20

21 **Transmission date:** April 9, 2013

**International Key Comparison(CCQM-K84) Report  
On purity analysis of methane**

Lab Information

Lab Name: National Institute of Metrology (NIM), China  
 Contact point: Dr. HAN Qiao and Dr. HU Shuguo  
 Email: [hanqiao@nim.ac.cn](mailto:hanqiao@nim.ac.cn), [hushg@nim.ac.cn](mailto:hushg@nim.ac.cn)  
 Tel.: +86-10-84252300 Fax.: +86-10-84252306  
 Date of Receiving the Comparison Cylinder: February, 2013  
 Cylinder No.: 0015220

Measurement and Result

Measurement #1

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	25/03/13	355.9	0.06%	2

Measurement #2

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	26/03/13	355.4	0.20%	5

Measurement #3

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	27/03/13	355.4	0.04%	3

Measurement #4

Component	Date (dd/mm/yy)	Result (nmol/mol)	Standard deviation (% relative)	Number of replicates
CO	28/03/13	355.4	0.05%	3

Results

Component	Result (nmol/mol)	Expanded Uncertainty (% relative)	Coverage factor**
CO	355.5	0.8%	2

\*\*The coverage

factor  $k=2$ (95% confidence level)

Method Description

1. Reference Method

CO was analyzed by GC-FID/Methanator (Agilent7890, Agilent, American) with a column

of molesieve 5A 80/100(9ft\*1/8inch\*2.0mm).

**GC conditions**

Oven temp: 110°C isotherm 4min

Sample loop: 5mL

Valve #1 load time: 0.1 ~ 0.4min

Carrier gas: N<sub>2</sub>

Carrier flow: 75 psi

Sample Flow: 400 mL/min

2. Calibration standard

**Cylinder No.: CAL017806**

**Preparation method**

All of the references we used were made by the gravimetric method according to ISO 6142-2006 by ourselves. The parent gases were filled into a 6-liter aluminum cylinder. At least, 10 g parent gas was filled into the cylinder. The cylinder was weighed before and after the filling using a balance with the sensitivity of 1 mg.

The concentration of reference gas was calculated according to the following equation.

$$x_i = \frac{\sum_{A=1}^P \left( \frac{x_{i,A} \cdot m_A}{\sum_{i=1}^n x_{i,A} \cdot M_i} \right)}{\sum_{A=1}^P \left( \frac{m_A}{\sum_{i=1}^n x_{i,A} \cdot M_i} \right)}$$

The uncertainty of reference gas included the contributions from gravimetric method and from stability. The uncertainty from stability was evaluated based on short-time and long-time testing before this comparison. The uncertainty from gravimetric method was calculated according to the following equation.

$$u^2(x_i) = \sum_{r=1}^q \left( \frac{\partial f_i}{\partial f_r} \right)^2 \cdot u^2(y_r) + 2 \sum_{r=1}^{q-1} \sum_{s=r+1}^q \left( \frac{\partial f_i}{\partial f_r} \right) \left( \frac{\partial f_i}{\partial f_s} \right) \cdot u(y_r, y_s)$$

Mass of parent gas filled, molecular weight and mole fraction of compound were the main sources of the uncertainty of gravimetric method.

**Purity analysis**

N<sub>2</sub>, O<sub>2</sub> and Ar were analyzed by GC-PDHID(Photo discharged helium ionization detector, Agilent 7890, American) with two columns of molesieve 5A (30m\*0.53mm\*15µm and 50m\*0.53mm\*15µm).

CO were analyzed by GC-PDHID and GC-FID with a column of Al<sub>2</sub>O<sub>3</sub>/KCl(50m\*0.53mm\*15µm).

**Uncertainty of calibration standard**

Component	Mole fraction 10 <sup>-6</sup> mol/mol	Expand Uncertainty(k=2)
-----------	---	----------------------------

		(10 <sup>-6</sup> mol/mol)
CO	0.3631	0.0024
H <sub>2</sub>	0.0500	0.0474
H <sub>2</sub> O	0.1298	0.1581
CO <sub>2</sub>	0.0100	0.0095
CH <sub>4</sub>	0.0200	0.0102
Ar	9981.0691	17.1687
O <sub>2</sub>	209081.9828	26.7492
N <sub>2</sub>	780936.3802	30.4943

### 3. Instrument calibration

When CO was analyzed, ‘A-B-A-B-A’ type calibration was used. That means the sample gas and our reference gas were measured in the order of Reference – Sample – Reference – Sample – Reference. The gas pressure at the sample loop of GC was controlled at almost same value during one analysis sequence. Single point calibration was used to calculate the concentration of target compound in sample cylinder.

### 4. Sampling handling

When package box including comparison cylinder arrived at the lab, it was in good state. Then the box was unpacked and the comparison cylinder was stored at room temperature. A SS regulator was connected to the cylinder.

During the analysis, the gas mixtures in both comparison cylinder and the reference cylinder, via regulators without pressure gauge, 1/8 inch stainless steel tube, were introduced into a 6-port valve. The pressure gauge and the mass flow meter were connected to the inlet of the 6-port valve to show the pressure and flow rate. The 6-port valve was driven by Nitrogen. The gas pressure before the sample loop was controlled at 0.1 MPa by regulator.

### Evaluation of measurement uncertainty

The contributions of measurement uncertainty were from reference gas, signal readings of the sample gas and reference gas, reproducibility in different days or groups.

$$u(c_{CCQM}) = \sqrt{u^2(c_{PRM}) + u^2(A_{CCQM}) + u^2(A_{PRM}) + u^2(f_{inter})}$$

Here,  $u$  means relative standard uncertainty.

$u(c_{CCQM})$ : Measurement uncertainty of concentration of the target component in the comparison sample gas cylinder.

1.  $u(A_{CCQM})$ : Uncertainty of signal reading of the sample gas from peak area on GC.

2.  $u(A_{PRM})$ : Uncertainty of signal reading of the reference gas from peak area on GC.

For the  $A_{CCQM}$  and  $A_{PRM}$ , the relative standard uncertainty could be calculated from the relative standard deviation (RSD) of the signal reading. The relative standard uncertainty is  $RSD/\sqrt{n}$ , where  $n$  is the number of signal reading.

3.  $u(c_{PRM})$ : Uncertainty of concentration of the reference gas, which was combined by the uncertainty from gravimetric method according to ISO 6142-2006 and the uncertainty from the stability of the reference gas.

4.  $u(f_{inter})$ : Uncertainty of reproducibility in different days or groups. The relative standard uncertainty  $f_{inter}$  was calculated from the relative standard deviation (RSD) of repeating test

in different days or groups. The relative standard uncertainty is  $RSD/\sqrt{n}$ , where n is the number of the repeating test.

Source of uncertainty	$u(c_{PRM})$	$u(A_{CCQM})$	$u(A_{PRM})$	$u(f_{inter})$
Relative standard uncertainty	0.33%	0.13%	0.13%	0.13%
Relative expanded uncertainty**	0.8%			

\*\*The coverage factor  $k=2$ (95% confidence level)

Report Form: carbon monoxide in synthetic air

Laboratory: National Physical Laboratory

Cylinder Number: D015215 (KC-9)

Measurement #1

Component	Date (dd/mm/yy)	Result (nmol/mol)	standard deviation (nmol/mol)	No. of replicates
CO	05/12/2012	355.0	1.4	3

Measurement#2

Component	Date (dd/mm/yy)	Result (nmol/mol)	standard deviation (nmol/mol)	No. of replicates
CO	05/12/2012	355.3	1.4	3

Measurement#3

Component	Date (dd/mm/yy)	Result (nmol/mol)	standard deviation (nmol/mol)	No. of replicates
CO	17/12/2012	354.1	1.4	4

Measurement#4

Component	Date (dd/mm/yy)	Result (nmol/mol)	standard deviation (nmol/mol)	No. of replicates
CO	11/01/2013	357.3	1.2	4

Results

Component	Date (dd/mm/yy)	Result (nmol/mol)	Expanded Uncertainty (nmol/mol)	Coverage Factor
CO	11/01/2013	355.4	7.0	2


## Details of the measurement method used

### Reference method

A Picarro G2401 Cavity ring-down spectrometer was used to measure the amount fraction of CO in the comparison cylinder.

### Calibration standards

Three reference standards were prepared at NPL in real scrubbed air from pure CO (99.9975%) and used in the analysis. All mixtures were prepared in BOC 10 litre cylinders with Spectraseal passivation. The scheme below shows the gravimetric dilutions with nominal CO amount fractions.



### Instrument calibration

The reference gas mixtures were prepared with CO amount fractions which lie within 2% of the comparison mixture. The analyser response to the matrix gas was recorded. A comparison of the reference mixture to the comparison mixture was achieved by measuring the analyser response to the reference mixture for a five minute period followed by the comparison mixture for the same time. This sequence was repeated a number of times (see results table). At the end of the experiment the analyser response to the matrix gas was recorded a second time. To minimise the effects from zero drift, a mean of the analyser response to the matrix gas before and after the experiment was used. The amount fraction of CO in the comparison mixture was determined using the amount fraction of the reference mixture and the ratio of the analyser response to the reference and unknown mixtures (both were corrected for the analyser response to matrix (zero) gas).

### Sample handling

Cylinders were maintained at a laboratory temperature of  $(20 \pm 3)$  °C throughout the period of analysis. Samples were introduced into the analyser at atmospheric pressure (excess flow was passed to vent) using a low volume gas regulator.

## Uncertainty

The estimated uncertainty for the measurement contains the following components:

- Purity analysis of CO and scrubbed real air
- Gravimetric preparation
- Analytical uncertainty from the spectrometer (drift and instrument noise)

## Uncertainty budget

The amount fraction of the comparison mixture,  $x_c$ , was determined using the following expression:

$$x_c = \frac{x_r(y_c - y_z)}{(y_r - y_z)}$$

where  $x_r$  is the amount fraction of the reference standard,  $y_c$ ,  $y_r$  and  $y_z$  are the analytical measurements of the comparison mixture, the reference standard and zero respectively. Both  $y_c$  and  $y_r$  are dominated by instrument repeatability. For the purposes of the uncertainty calculation, the equation above represents a situation where repeatability of the measurement takes into account any drift over the measurement period.

The uncertainty in the amount fraction of the comparison cylinder was determined by adding the four components in quadrature. The table which follows details the uncertainty analysis.

quantity	units	example value	standard uncertainty	sensitivity coefficient	uncertainty contribution	uncertainty type	distribution
$x_r$	nmol/mol	353.5	2.0	1.008	2.012	A	normal
$y_z$	nmol/mol	3.1	1.5	0.008	0.011	A	normal
$y_r$	nmol/mol	353.2	2.0	-1.017	-2.035	A	normal
$y_c$	nmol/mol	355.8	2.0	1.010	2.020	A	normal
$x_c$	nmol/mol	356.1					
$u(x_c)$	nmol/mol	3.5					
$U(x_c)$	nmol/mol	7.0					



**Report Form Carbon monoxide in synthetic air  
Key Comparison CCQM-K84**

**Laboratory Information.**

**Institute:** NOAA/OAR/ESRL/GMD

**Address:** 325 Broadway street  
Mail Stop R.GMD1  
Boulder, Colorado USA 80305

**Contact:** Brad Hall, Paul Noveilli

**Tel:** +1 303.497.7011

**Email:** [Bradley.Hall@noaa.gov](mailto:Bradley.Hall@noaa.gov)

**Serial no.** DO15283

**Reference Method.**

**Analytical methods:** The CO mole fraction in cylinder DO15283 (Table 1) was determined using enhanced off-axis integrated-cavity output spectroscopy (ICOS). A CO/N<sub>2</sub>O analyzer from Los Gatos Research Inc. (Mountain View California, USA, model 907-0015, serial no. 09-007) was used. The instrument was upgraded with software for Los Gatos to improve resolution of the instrument signal.

**Reference Gases:** Measurements are referenced to the WMO CO X2004 scale (WMO, 2010) and reported in nmol CO per mol dry air. The primary standards which define the scale were prepared by a gravimetric method. Secondary standards were assigned mole fractions by calibration against two sets of primary standards. Uncertainty in the scale is obtained from measured differences in mole fraction results for standards calibrated using independent sets of primary standards.

**Calibration procedure:** Instrument response was modeled as a nonlinear function. Response curves were obtained from measurement signals of nine secondary standards covering 25 to 500 nmol mol<sup>-1</sup>. The one-second instrument measurement signals approach a Gaussian distribution and were averaged over 90s. The mean instrument response of each standard, obtained from eight individual determinations of 90 1 Hz measurements, were fit with an orthogonal polynomial (n=2) to define instrument response.

Sample mole fractions are obtained from sample measurement response and the response curve. Four calibration events were conducted over an 8-week period. Each event consisted of 15 determinations of sample response (mean of 90 1 Hz signals) translated to mole fraction. The value reported here represents the mean of four calibrations events.

**Uncertainty of Results:** The uncertainty of measurement results is estimated by a statistical approach to yield a combined standard uncertainty expanded to the 95% confidence level (coverage factor = 2). The combined uncertainty includes scale and measurement uncertainties obtained by propagation in quadrature of the uncertainties in the WMO CO X2004 reference scale, the response curve and the repeatability of multiple calibration events of the sample.

Table1. Nominal composition of mixture

Component	X (nominal value)
Carbon Monoxide [nmol/mol]	300
Argon [% mol/mol]	0.93
Oxygen [% mol/mol]	21
Nitrogen (5 nmol/mol)	balance

Table 2. Uncertainty components

Component	Mean Uncertainty 1 $\sigma$ [nmol/mol]	Fractional uncertainty (%)
<i>Gravimetric Standards</i>		
Weighing uncertainty	1.21	0.35
CO in dilution gas	2	0.58
<i>Analytical</i>		
Response curve	0.24	0.07
Reproducibility	0.17	0.05
<i>Total Combined</i>	2.35	0.68

## Results

### Measurement #1

Component	Date (dd/mm/yy)	Result (nmol /mol)	Standard Deviation (% relative to Result)	Number of replicates
CO	24/09/12	346.06	3.22e-4	14

### Measurement #2

Component	Date (dd/mm/yy)	Result (nmol /mol)	Standard Deviation (% relative to Result)	Number of replicates
CO	23/10/12	346.13	3.47e-4	14

### Measurement #3

Component	Date (dd/mm/yy)	Result (nmol /mol)	Standard Deviation (% relative to Result)	Number of replicates
CO	21/11/12	346.36	1.76e-4	14

### Measurement #4

Component	Date (dd/mm/yy)	Result (nmol /mol)	Standard Deviation (% relative to Result)	Number of replicates
CO	03/12/12	347.01	4.62e-4	14

## Result

Component	Result (nmol /mol)	Expanded Uncertainty (nmol / mol)	Coverage Factor
CO	346.0	4.7	2

**Note:** CO contained in cylinders of the size and material comparable to those used in this experiment has been known to change with time.

Appendix A.

## Report Form Carbon monoxide in synthetic air

Laboratory name: Korea Research Institute of Standards and Science (KRISS)

Cylinder number: D015286

### Measurement #1

Component	Date (dd/mm/yy)	Result (nmol /mol)	Standard deviation (nmol /mol)	number of replicates
CO	26/07/12	353.16	0.50	4
	4/09/12	353.34	0.60	4
	4/09/12	353.22	0.64	3
	5/09/12	353.18	0.50	4
	14/09/12	353.36	0.74	4

### Results

Component	Result (nmol/mol)	Expanded Uncertainty (nmol /mol)	Coverage factor <sup>3</sup>
CO	353.25	1.06	2

### Method Description Forms

#### Details of the measurement method used:

##### **Analysis method:**

Carbon monoxide concentration in synthetic air has been quantified using gas chromatograph thermal conductivity detector with Methanator (GC-TCD/Methanator). Figure 1 shows an analytical condition of the analyzer and its chromatogram.


To achieve analytical interval of  $\pm 0.1\%$  (standard deviation) the instrument drift and standard deviation of the response were controlled carefully. The cylinder D015286 were analyzed against the primary reference mixture of D985725 (prepared in July, 2012).

---

<sup>3</sup> The coverage factor shall be based on approximately 95% confidence.

<Analytical condition>

Detector : FID/Methanator  
Detector temp. : 275 °C  
H<sub>2</sub> : 95 mL/min, Air : 350 mL/min  
Oven temp. : 80 °C  
Column : Restek Molesieve 5A, 80/100 16ft\*1/8" SS  
Carrier gas : N<sub>2</sub>, 95 psi  
Sample loop vol. : 10 mL with restrictor  
Sample flow : 80 mL/min  
matrix effect (no consideration), valve(0.1/1.1)



**Instrument calibration:**

Instrument calibration is performed using KRISS primary standard mixtures. One point calibration was done with a cylinder of nominal value ~ 350 nmol/mol which was very close to the target cylinder.

**Sample handling:**

The sample cylinder had put in the laboratory with room temperature for several days after preparation. Each cylinder was equipped with a stainless steel pressure regulator that was purged more than 7 times after connection to the analysis line. Samples were transferred to sample loop at flow rate of 80 ml/min using mass-flow controller.

**Calibration standards:**

Preparation method

5 primary standard mixtures were used for the determination of carbon monoxide in synthetic air. The standards were prepared from pure carbon monoxide, pure nitrogen, and pure oxygen in accordance with ISO6142:2001 (Gas analysis-Preparation of calibration gases-Gravimetric method. Pure carbon monoxide was diluted by 4 step and purity analysis for every pure gases were done. Table 1 shows gravimetric value and expanded uncertainty of the calibration standards. They agreed within 0.1 % as shown in Figure 1.

Table 1. Gravimetric value and expanded uncertainty in calibration standards

Cylinder number	Gravimetric Value (nmol/mol)	Expanded uncertainty [k=2] (nmol/mol)
D905128	351.08	0.72
D905126	347.61	0.69
D929208	348.75	0.70
D985725	341.52	0.68
D985730	342.95	0.68




Figure 5. Consistency among primary standard mixtures

### Purity analysis

The impurities of carbon monoxide, nitrogen, and oxygen were determined by analytical methods and the amount of the major component is conventionally determined from the following equation,

$$x_{pure} = 1 - \sum_{i=1}^N x_i$$

where

$x_i$ : the mole fraction of impurity  $i$ , determined by analysis;

$N$ : the number of impurities likely to be present in the final mixture;

$x_{pure}$ : the mole fraction “purity” of the “pure” parent gas.

Table 2-4 shows summarized results of purity analyses for CO, N<sub>2</sub>, and O<sub>2</sub>. The purity results of them were considered in gravimetric preparation, CO in Oxygen was added to the gravimetric value as well as the uncertainty. Total uncertainty of CO was calculated with GUM program. For purity analysis GC-AED, TCD, and PDD were applied. High value of CO in Oxygen acts as a major contributor of uncertainty during preparation.

Table 2. Results of Purity analysis of carbon monoxide (QA8272, 50L Al)

component	analytical conc. (umol/mol)	distribution		applied conc. (umol/mol)	standard uncertainty (umol/mol)	f*f
H2	< 0.26	rectangular	1.732	0.13	0.075	0.005633
H2O	<1.0	rectangular	1.732	0.5	0.289	0.083333
CH4	<0.08	rectangular	1.732	0.04	0.023	0.000533
CO2	<1.02	rectangular	1.732	0.51	0.294	0.086700
THC	<1.0	rectangular	1.732	0.5	0.289	0.083333
N2	4.13	normal	0.2	4.13	0.413	0.170569
O2+Ar	0.93	normal	0.2	0.93	0.093	0.008649
impurities				6.740	0.662	0.438751
<b>CO</b>				999993.260	1.325	k=2

Table 3. Results of Purity analysis of Nitrogen

component	analytical conc. (umol/mol)	distribution		applied conc. (umol/mol)	standard uncertainty (umol/mol)	f*f
H2	< 0.5	rectangular	1.732	0.25	0.144	0.020833
H2O	1.2	normal	0.2	1.2	0.120	0.014400
CO	<0.002	rectangular	1.732	0.001	0.001	0.000000
CH4	< 0.001	rectangular	1.732	0.0005	0.000	0.000000
CO2	< 0.01	rectangular	1.732	0.005	0.003	0.000008
THC	< 0.5	rectangular	1.732	0.25	0.144	0.020833
Ar	< 1.0	rectangular	1.732	0.5	0.289	0.083333
O2	0.35	normal	0.2	0.35	0.035	0.001225
Ne	< 1.0	rectangular	1.732	0.5	0.289	0.083333
impurities				3.057	0.473	0.223967
<b>N2</b>				999996.944	0.947	k=2

Table 4. Results of Purity analysis of Oxygen

component	analytical conc. (umol/mol)	distribution		applied conc. (umol/mol)	standard uncertainty (umol/mol)	f*f
H2	< 0.1	rectangular	1.732	0.05	0.029	0.000833
H2O	1.54	normal	0.2	1.54	0.154	0.023716
CO	0.00685	normal	0.5	0.00685	0.00171	0.00000293
CH4	< 0.1	rectangular	1.732	0.05	0.029	0.000833
CO2	0.22	normal	0.2	0.22	0.022	0.000484
THC	< 0.3	rectangular	1.732	0.15	0.087	0.007500
Ar	< 1.0	rectangular	1.732	0.5	0.289	0.083333
N2	5.84	normal	0.2	5.84	0.584	0.341056
impurities				8.357	0.677	0.457759
<b>O2</b>				999991.643	1.353	k=2

### Uncertainty:

The uncertainty used for the calibration mixtures contains all sources of gravimetric preparation. Uncertainty for stability is not included because no instability has been detected. An analysis uncertainty is calculated based on repeatability and drift of analyzer of the acquired area.

### Detailed uncertainty budget:

Please include a list of the uncertainty contributions, the estimate of the standard uncertainty, probability distributions, sensitivity coefficients, etc.

Typical evaluation of the measurement uncertainty for CO:

Quantity $X_i$	Estimate $x_i$	Evaluation Type (A or B)	Distribution	Standard uncertainty $u(x_i)$ [nmol/mol]	Sensitivity coefficient <i>Rel. <math>u(x_i)</math></i> [%]	Contribution $u_i(y)$
References		A	Gaussian	0.35	0.1	
Sample		A	Gaussian	0.35	0.1	
References prepared grav.		A	Gaussian	0.42	0.12	
Combined standard uncertainty				0.65	0.18	