

Table 4. Equipment and source of UTC(k) of the laboratories contributing to TAI in 2016

Ind. Cs: industrial caesium standard
 Ind. Rb: industrial rubidium standard
 Lab. Cs: laboratory caesium standard
 Lab. Rb: laboratory rubidium standard
 H-maser: hydrogen maser
 SF: single frequency receiver
 DF: dual frequency receiver
 * means 'yes'

Lab <i>k</i>	Equipment	Source of UTC(<i>k</i>) (1)	TA(<i>k</i>)	UTC <i>r</i>	Time Links			
					GPS		GLONASS	Two-Way
					SF	DF		
AOS	3 Ind. Cs 2 H-masers	1 H-maser + microphase-stepper (2)	* (19)	*	*	*	*	*
APL	3 Ind. Cs 3 H-masers	1 H-maser + frequency synthesizer steered to UTC(APL)			*	*		
AUS	5 Ind. Cs 2 H-masers	1 Cs		*	*	*	*	*
BEV	2 Ind. Cs 1 H-maser	1 Cs		*	*	*	*	
BIM	2 Ind. Cs	1 Cs			*	*		
BIRM	2 Ind. Cs 3 H-masers	1 H-maser + microphase-stepper			*			
BY (a)	6 H-masers	3-4 H-masers + microphase-stepper			*		*	
CAO	2 Ind. Cs	1 Cs			*	*	*	
CH	3 Ind. Cs (3) 2 H-maser	1 H-maser (3) + frequency synthesizer steered to UTC(CH.P)	*	*		*		*
CNES	8 Ind. Cs (4) 3 H-maser	1 H-maser (4) + microphase-stepper				*		

Table 4. Equipment and source of UTC(k) of the laboratories contributing to TAI in 2016 (Cont.)

Ind. Cs: industrial caesium standard
 Ind. Rb: industrial rubidium standard
 Lab. Cs: laboratory caesium standard
 Lab. Rb: laboratory rubidium standard
 H-maser: hydrogen maser
 SF: single frequency receiver
 DF: dual frequency receiver
 * means 'yes'

Lab <i>k</i>	Equipment	Source of UTC(<i>k</i>) (1)	TA(<i>k</i>)	UTC <i>r</i>	Time Links			
					GPS		GLONASS	Two-Way
					SF	DF		
CNM (a)	4 Ind. Cs 2 H-maser	1 H-maser + microphase-stepper	*	*		*	*	
CNMP	3 Ind. Cs	1 Cs + frequency offset generator		*	*	*	*	
DFNT	2 Ind. Cs	1 Cs				*	*	
DLR	3 Ind. Cs 3 H-maser	1 Cs		*		*	*	
DMDM	2 Ind. Cs	1 Cs + microphase-stepper		*	*	*		
DTAG	3 Ind. Cs	1 Cs		*		*		
EIM	2 Ind. Cs	1 Cs			*			
ESTC	3 Ind. Cs 3 H-masers	1 H-maser + microphase-stepper		*		*		
HKO	2 Ind. Cs	1 Cs				*	*	
IFAG	5 Ind. Cs 2 H-masers	1 Cs + microphase-stepper		*		*		

Table 4. Equipment and source of UTC(k) of the laboratories contributing to TAI in 2016 (Cont.)

Ind. Cs: industrial caesium standard
 Ind. Rb: industrial rubidium standard
 Lab. Cs: laboratory caesium standard
 Lab. Rb: laboratory rubidium standard
 H-maser: hydrogen maser
 SF: single frequency receiver
 DF: dual frequency receiver
 * means 'yes'

Lab <i>k</i>	Equipment	Source of UTC(<i>k</i>) (1)	TA(<i>k</i>)	UTC <i>r</i>	Time Links			
					GPS		GLONASS	Two-Way
					SF	DF		
IGNA	1 Ind. Cs	1 Cs			*			
IMBH	2 Ind. Cs	1 Cs		*		*	*	
INCP	1 Ind. Cs	1 Cs			*			
INPL	2 Ind. Cs	1 Cs				*	*	
INTI	1 Ind. Cs	1 Cs		*	*			
INXE	3 Ind. Cs 1 Ind. Rb 1 Lab. Cs	1 Cs + microphase-stepper		*	*	*		
IT	6 Ind. Cs 4 H-masers 2 Lab. Cs	1 H-maser + microphase-stepper		*		*		*
JATC	(5)	1 Cs + microphase-stepper	*					
JV	3 Ind. Cs 1 H-maser	1 H-maser + microphase-stepper		*		*	*	
KEBS	3 Ind. Cs	1 Cs + reference generator				*	*	

Table 4. Equipment and source of UTC(k) of the laboratories contributing to TAI in 2016 (Cont.)

Ind. Cs: industrial caesium standard
 Ind. Rb: industrial rubidium standard
 Lab. Cs: laboratory caesium standard
 Lab. Rb: laboratory rubidium standard
 H-maser: hydrogen maser
 SF: single frequency receiver
 DF: dual frequency receiver
 * means 'yes'

Lab <i>k</i>	Equipment	Source of UTC(<i>k</i>) (1)	TA(<i>k</i>)	UTC <i>r</i>	Time Links			
					GPS		GLONASS	Two-Way
					SF	DF		
KIM (a)	2 Ind. Cs	1 Cs				*	*	
KRIS (a)	5 Ind. Cs 4 H-masers	1 H-maser + microphase-stepper	*	*	*	*	*	*
KZ	5 Ind. Cs (6)	1 Cs + microphase-stepper				*	*	
LT	2 Ind. Cs	1 Cs		*	*			
MASM	1 Ind. Cs	1 Cs + time/frequency steering				*	*	
MBM	1 Ind. Cs	1 Cs			*			
MIKE	1 Ind. Cs 4 H-masers	1 H-maser + microphase-stepper		*	*	*		
MKEH	1 Ind. Cs	1 Cs			*			
MSL	2 Ind. Cs	1 Cs + microphase-stepper		*		*		
MTC	11 Ind. Cs	1 Cs		*	*	*	*	

Table 4. Equipment and source of UTC(k) of the laboratories contributing to TAI in 2016 (Cont.)

Ind. Cs: industrial caesium standard
 Ind. Rb: industrial rubidium standard
 Lab. Cs: laboratory caesium standard
 Lab. Rb: laboratory rubidium standard
 H-maser: hydrogen maser
 SF: single frequency receiver
 DF: dual frequency receiver
 * means 'yes'

Lab <i>k</i>	Equipment	Source of UTC(<i>k</i>) (1)	TA(<i>k</i>)	UTC <i>r</i>	Time Links			
					GPS		GLONASS	Two-Way
					SF	DF		
NAO	4 Ind. Cs 1 H-maser	1 Cs + microphase-stepper		*	*			
NICT	33 Ind. Cs 9 H-masers (7) 1 Lab. Cs	1 H-maser (8) + microphase-stepper	* (9)	*	*	*		*
NIM	7 Ind. Cs 6 H-masers 1 Lab. Cs	1 H-maser + microphase-stepper		*	*	*	*	*
NIMB	2 Ind. Cs	1 Cs		*	*	*		
NIMT	2 Ind. Cs	1 Cs + microphase-stepper		*	*	*		
NIS	2 Ind. Cs	1 Cs			*	*	*	
NIST	2 Lab. Cs 15 Ind. Cs 13 H-masers	6 Cs 8 H-masers + microphase-stepper	*	*		*		*
NMIJ	3 Ind. Cs 1 Lab. Cs 4 H-masers	1 H-maser + microphase-stepper		*	*	*		*
NMLS (a)	2 Ind. Cs	1 Cs		*		*		
NPL	2 Ind. Cs 5 H-masers	1 H-maser			*	*		*

Table 4. Equipment and source of UTC(k) of the laboratories contributing to TAI in 2016 (Cont.)

Ind. Cs: industrial caesium standard
 Ind. Rb: industrial rubidium standard
 Lab. Cs: laboratory caesium standard
 Lab. Rb: laboratory rubidium standard
 H-maser: hydrogen maser
 SF: single frequency receiver
 DF: dual frequency receiver
 * means 'yes'

Lab <i>k</i>	Equipment	Source of UTC(<i>k</i>) (1)	TA(<i>k</i>)	UTC <i>r</i>	Time Links			
					GPS		GLONASS	Two-Way
					SF	DF		
NPLI	5 Ind. Cs 1 H-maser	1 H-maser + microphase-stepper		*		*		*
NRC	6 Ind. Cs (10) 1 H-masers	1 Cs + microphase-stepper	*	*		*		
NRL	9 Ind. Cs 9 H-masers	1 H-maser + frequency synthesizer steered to UTC(NRL)		*		*		
NTSC	24 Ind. Cs 4 H-masers	1 H-maser + microphase-stepper	*	*	*	*		*
ONBA	2 Ind. Cs	1 Cs			*			
ONRJ	7 Ind. Cs 2 H-masers	7 Cs 2 H-masers + frequency offset generator	*	*		*	*	
OP	5 Ind. Cs 3 Lab. Cs 1 Lab. Rb 4 H-masers	1 H-maser (12) + microphase-stepper	*	*	*	*	*	*
ORB	4 Ind. Cs 1 H-masers	1 H-maser or 1 Cs (14) + femtostepper		*		*	*	
PL	13 Ind. Cs 4 H-masers	1 Cs (15) + femtostepper	*	*	*	*	*	
PTB	3 Ind. Cs 4 Lab. Cs (17) 4 H-masers	1 H-maser (18) + microphase-stepper	*	*	*	*	*	*

Table 4. Equipment and source of UTC(k) of the laboratories contributing to TAI in 2016 (Cont.)

Ind. Cs: industrial caesium standard
 Ind. Rb: industrial rubidium standard
 Lab. Cs: laboratory caesium standard
 Lab. Rb: laboratory rubidium standard
 H-maser: hydrogen maser
 SF: single frequency receiver
 DF: dual frequency receiver
 * means 'yes'

Lab <i>k</i>	Equipment	Source of UTC(<i>k</i>) (1)	TA(<i>k</i>)	UTC <i>r</i>	Time Links			
					GPS		GLONASS	Two-Way
					SF	DF		
ROA	6 Ind. Cs (20) 2 H-maser	1 H-maser (21) + frequency synthesizer steered to UTC(ROA)		*		*	*	*
SASO	5 Ind. Cs	1 Cs		*		*	*	
SCL	2 Ind. Cs	1 Cs + microphase-stepper		*	*			
SG	5 Ind. Cs 1 H-maser	1 H-maser + microphase-stepper	*	*		*	*	
SIQ	1 Ind. Cs	1 Cs				*	*	
SMD	3 Ind. Cs 1 H-maser	1 Cs + microphase-stepper		*		*	*	
SMU	1 Ind. Cs	1 Cs + output frequency steering			*	*	*	
SP	19 Ind. Cs (22) 9 H-masers	1 H-maser + microphase-stepper		*		*	*	*
SU	2 Lab. Cs (23) 9-12 H-masers	6-9 H-masers (24)	* (25)	*		*	*	*
TL	10 Ind. Cs 4 H-masers	1 H-maser + microphase-stepper	* (26)	*		*		*

Table 4. Equipment and source of UTC(k) of the laboratories contributing to TAI in 2016 (Cont.)

Ind. Cs: industrial caesium standard
 Ind. Rb: industrial rubidium standard
 Lab. Cs: laboratory caesium standard
 Lab. Rb: laboratory rubidium standard
 H-maser: hydrogen maser
 SF: single frequency receiver
 DF: dual frequency receiver
 * means 'yes'

Lab <i>k</i>	Equipment	Source of UTC(<i>k</i>) (1)	TA(<i>k</i>)	UTC <i>r</i>	Time Links			
					GPS		GLONASS	Two-Way
					SF	DF		
TP	5 Ind. Cs	1 Cs + output frequency steering				*	*	
UA (a)	1 Ind. Cs 3 H-masers	3 H-masers + microphase-stepper			*	*	*	
UME	5 Ind. Cs	1 Cs		*		*	*	
USNO (a)	81 Ind. Cs 33 H-masers 6 Lab. Rb	1 H-maser ⁽²⁷⁾ + frequency synthesizer steered to UTC(USNO)	* (27)	*	*	*		*
VMI	3 Ind. Cs	1 Cs + microphase-stepper				*		
VSL	4 Ind. Cs	1 Cs + microphase-stepper		*		*		*
ZA	6 Ind. Cs 1 H-maser	1 H-maser				*	*	

Notes

- (a) Information based on the Annual Report for 2015, not confirmed by the laboratory.
- (1) When several clocks are indicated as a source of UTC(*k*), laboratory *k* computes a software clock, steered to UTC. Often a physical realization of UTC(*k*) is obtained using a Cs clock and a micro-phase-stepper.
- (2) AOS The UTC(AOS) is formed technically using 1 hydrogen maser and microstepper, it is steered using TA(PL) data as a reference.
TA(PL) laboratories are linked via MC GPS-CV and/or two-directional optical fibre connections. Optical Fibre Link UTC(AOS)-UTC(PL) is 420 km long.
- (3) CH All the standards are located in Bern at METAS (Swiss Federal Institute of Metrology). Since November 2007, UTC(CH) is defined in real time by a hydrogen maser steered to the paper time scale UTC(CH.P) which is defined as a weighted average of all the clocks, steered to UTC.
TA(CH) is also a weighted average of all the clocks, but free running.
- (4) CNES All the standards are located in Toulouse at CNES (French Space Agency). UTC(CNES) is defined in real time by a H-Maser steered to an ensemble of industrial high-performance Cs clocks.
UTC(CNES) is steered monthly on UTC.
- (5) JATC The standards are located at National Time Service Centre (NTSC). The link between UTC(JATC) and UTC(NTSC) is obtained by internal connection.
- (6) KZ The standards are located as follows:
- | | |
|---|------|
| *Kazakhstan Institute for Metrology (Astana) | 4 Cs |
| *South-Kazakhstan branch of Kazakhstan Institute for Metrology (Almaty) | 1 Cs |
- (7) NICT The standards are located as follows:
- | | |
|---|-------------------|
| * Koganei Headquarters | 20 Cs, 7 H-masers |
| * Ohtakadoya-yama LF station | 4 Cs |
| * Hagane-yama LF station | 5 Cs |
| * Advanced ICT Research Institute in Kobe | 5 Cs, 2 H-masers |
- (8) NICT The NICT atomic timescale TA(NICT) is computed from the weighted average of 18 commercial Cs clocks at the Koganei HQ.
- (9) NICT UTC(NICT) is generated from the output of a hydrogen maser, steered to TA(NICT) regularly, and steered to UTC if necessary.
- (10) NRC The standards are located as follows:
- | | |
|--|-----------------|
| * Measurement Science and Standards (Ottawa) | 4 Cs, 1 H-maser |
| * CHU Time signal radio station (Ottawa) | 2 Cs |
- (11) ONRJ The Brazilian atomic time scale TA(ONRJ) is computed by the National Observatory Time Service Division in Rio de Janeiro with data from 7 industrial caesium clocks and 2 hydrogen masers.
- (12) OP Since MJD 56218 UTC(OP) is based on the output signal of a H-maser frequency steered towards UTC using the LNE-SYRTE fountains calibrations.

Notes (Cont.)

- (13) OP The French atomic time scale TA(F) is computed by the LNE-SYRTE with data from up to 21 industrial caesium clocks in 2016 located as follows :
- | | |
|---|------|
| * Centre Electronique de l'Armement (CELAR, Rennes) | 1 Cs |
| * Centre National d'Etudes Spatiales (CNES, Toulouse) | 3 Cs |
| * France Telecom Recherche et Developpement (Lannion) | 2 Cs |
| * Observatoire de la Côte d'Azur (OCA, Grasse) | 2 Cs |
| * Observatoire de Paris (LNE-SYRTE, Paris) | 5 Cs |
| * Observatoire de Besançon (OB, Besançon) | 3 Cs |
| * Direction des Constructions Navales (DCN, Brest) | 4 Cs |
| * Spectracom, Orolia (Les Ulis) | 1 Cs |
- All laboratories are linked via GPS receivers. The TA(F) frequency is steered using the LNE-SYRTE PFS data. The difference TA(F) – UTC(OP) is published in the OP Time Service Bulletin.
- (14) ORB The source of UTC(ORB) is generated by a Cs clock since July 2013.
- (15) PL The Polish official timescale UTC(PL) is maintained by the GUM.
- (16) PL The Polish atomic timescale TA(PL) is computed by the AOS and GUM with data from 15 caesium clocks and 4 hydrogen masers located as follows:
- | | |
|---|------------------|
| * Central Office of Measures (GUM, Warsaw) | 3 Cs, 1 H-maser |
| * Astrogeodynamical Observatory, Space Research Center P.A.S. (AOS, Borowiec) | 2 Cs, 2 H-masers |
| * National Institute of Telecommunications (IŁ, Warsaw) | 2 Cs |
| * Polish Telecom (Orange Polska S.A., Warsaw) | 3 Cs |
| * Military Primary Standards Laboratory (CWOM, Warsaw and Poznan) | 3 Cs |
| * Poznan Supercomputing and Networking Center (PSNC, Poznan) | 1 H-maser |
- and additionally
- | | |
|---|------|
| * Time and Frequency Standard Laboratory of the Center for Physical Science and Technology (FTMC), a guest laboratory from Lithuania (LT, Vilnius, Lithuania) | 2 Cs |
|---|------|
- All laboratories are linked via MC GPS-CV and/or two-directional optical fibre connections.
- (17) PTB The laboratory Cs, PTB CS1 and PTB CS2 are operated continuously as clocks. PTB CSF1 and CSF2 are fountain frequency standards using laser cooled caesium atoms. Both are intermittently operated as frequency standards. Contributions to TAI are made through comparisons with one of PTB's hydrogen masers.
- (18) PTB UTC(PTB) is based on the output of an active hydrogen maser steered in frequency since MJD 55224 (February 2010).
- (19) PTB Since MJD 56079 0:00 UTC TA(PTB) has been generated from an active hydrogen maser, steered in frequency so as to follow PTB caesium fountains as close as possible. The deviation d between the fountains and the TAI second is not taken into account. TAI-TA(PTB) got an initial arbitrary offset from TAI without continuity to the data reported in previous months.

Notes (Cont.)

- (20) ROA The standards are located as follows:
- | | |
|---|-----------------|
| * Real Observatorio de la Armada en San Fernando | 5 Cs, 1 H-maser |
| * Centro Español de Metrología | 1 Cs |
| * Added in October 2016, not yet declared to the BIPM | 1 H-maser |
- (21) ROA Since March 2009, UTC(ROA) is defined in real time by a hydrogen maser, steered to the paper time scale UTC(ROA), which is defined as a weighted average of all the clocks, steered to UTC.
- (22) SP The standards are located as follows (at the end of 2016):
- | | |
|---|------------------|
| * SP Technical Research Institute of Sweden (SP, Borås) | 4 Cs, 2 H-masers |
| * SP Technical Research Institute of Sweden (SP, Stockholm) | 6 Cs, 2 H-masers |
| * STUPI AB (Stockholm) | 8 Cs, 3 H-masers |
| * Onsala Space Observatory (Onsala) | 1 Cs, 2 H-masers |
- (23) SU CsFO1 and CsFO2 are fountain frequency standards using laser cooled caesium atoms. CsFO2 operated as frequency standard almost regularly and contributed to TAI.
- (24) SU Laboratory computes UTC(SU) as a software clock, steered to UTC.
- (25) SU TA(SU) is generated from an ensemble of active hydrogen masers, software steered in frequency so as to follow SU caesium fountains as close as possible. The deviation d between the fountains and the TAI second is not taken into account. TAI-TA(SU) has an initial arbitrary offset from TAI.
- (26) TL TA(TL) is generated from a 10-caesium-clock ensemble.
- (27) USNO The time scales A.1(MEAN) and UTC(USNO) are computed by USNO. They are determined by a weighted average of Cs clocks, hydrogen masers, and rubidium fountains located at the USNO. A.1(MEAN) is a free atomic time scale, while UTC(USNO) is steered to UTC. Included in the total number of USNO atomic standards are the clocks located at the USNO Alternate Master Clock in Colorado Springs, CO.